

Laboratory Control Sample Summary

Client:	Weston Solutions	SDG No.:	Q1569
Project:	Fort Meade MD Tipton Airfield Parcel RI - 0111169	Run No.:	LB135044

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167142BS							
Cyanide	mg/Kg	5	4.70		94	1	85-115	03/14/2025

Laboratory Control Sample Summary

Client:	Weston Solutions	SDG No.:	Q1569
Project:	Fort Meade MD Tipton Airfield Parcel RI - 0111169	Run No.:	LB135093

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167200BS							
Sulfide	mg/Kg	250	216		86	1	80-120	03/19/2025

Laboratory Control Sample Summary

Client:	Weston Solutions	SDG No.:	Q1569
Project:	Fort Meade MD Tipton Airfield Parcel RI - 0111169	Run No.:	LB135124

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167244BS							
Cyanide	mg/Kg	5	4.50		90	1	85-115	03/21/2025