



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Weston Solutions SDG No.: Q1569
Contract: WEST04 Lab Code: CHEM Case No.: Q1569 SAS No.: Q1569
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	19.6	20.0	98	40 - 160	P	03/17/2025	14:47	LB135063
	Barium	92.8	100	93	40 - 160	P	03/17/2025	14:47	LB135063
	Cadmium	5.88	6.0	98	40 - 160	P	03/17/2025	14:47	LB135063
	Chromium	10.3	10.0	103	40 - 160	P	03/17/2025	14:47	LB135063
	Lead	12.6	12.0	105	40 - 160	P	03/17/2025	14:47	LB135063
	Selenium	18.9	20.0	94	40 - 160	P	03/17/2025	14:47	LB135063
	Silver	10.3	10.0	103	40 - 160	P	03/17/2025	14:47	LB135063
CRA	Mercury	0.16	0.2	81	40 - 160	CV	03/18/2025	10:06	LB135064
CRI01	Arsenic	22.6	20.0	113	40 - 160	P	03/18/2025	10:49	LB135083
	Barium	94.4	100	94	40 - 160	P	03/18/2025	10:49	LB135083
	Cadmium	5.91	6.0	98	40 - 160	P	03/18/2025	10:49	LB135083
	Chromium	10.8	10.0	108	40 - 160	P	03/18/2025	10:49	LB135083
	Lead	11.3	12.0	94	40 - 160	P	03/18/2025	10:49	LB135083
	Selenium	16.6	20.0	83	40 - 160	P	03/18/2025	10:49	LB135083
	Silver	10.7	10.0	107	40 - 160	P	03/18/2025	10:49	LB135083
CRI01	Arsenic	21.2	20.0	106	40 - 160	P	03/25/2025	10:22	LB135182
	Barium	91.1	100	91	40 - 160	P	03/25/2025	10:22	LB135182
	Cadmium	5.71	6.0	95	40 - 160	P	03/25/2025	10:22	LB135182
	Chromium	9.81	10.0	98	40 - 160	P	03/25/2025	10:22	LB135182
	Lead	11.8	12.0	98	40 - 160	P	03/25/2025	10:22	LB135182
	Selenium	21.9	20.0	110	40 - 160	P	03/25/2025	10:22	LB135182
	Silver	11.2	10.0	112	40 - 160	P	03/25/2025	10:22	LB135182