

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Weston Solutions **SDG No.:** Q1569
Contract: WEST04 **Lab Code:** CHEM **Case No.:** Q1569 **SAS No.:** Q1569
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	4.68			-20	20	03/17/2025	14:51	LB135063
	Barium	3.10	6.0	52	-94	106	03/17/2025	14:51	LB135063
	Cadmium	-4.02	1.0	402	-5	7	03/17/2025	14:51	LB135063
	Chromium	60.2	52.0	116	42	62	03/17/2025	14:51	LB135063
	Lead	7.22			-12	12	03/17/2025	14:51	LB135063
	Selenium	-0.91			-20	20	03/17/2025	14:51	LB135063
	Silver	-2.37			-10	10	03/17/2025	14:51	LB135063
ICSAB01	Arsenic	95.9	104	92	88.4	120	03/17/2025	15:02	LB135063
	Barium	491	537	91	437	637	03/17/2025	15:02	LB135063
	Cadmium	1030	972	106	826	1120	03/17/2025	15:02	LB135063
	Chromium	588	542	108	460	624	03/17/2025	15:02	LB135063
	Lead	58.7	49.0	120	37	61	03/17/2025	15:02	LB135063
	Selenium	57.1	46.0	124	26	66	03/17/2025	15:02	LB135063
	Silver	224	201	111	170	232	03/17/2025	15:02	LB135063
ICSA01	Arsenic	3.99			-20	20	03/18/2025	10:54	LB135083
	Barium	5.93	6.0	99	-94	106	03/18/2025	10:54	LB135083
	Cadmium	-4.36	1.0	436	-5	7	03/18/2025	10:54	LB135083
	Chromium	60.3	52.0	116	42	62	03/18/2025	10:54	LB135083
	Lead	5.17			-12	12	03/18/2025	10:54	LB135083
	Selenium	-12.3			-20	20	03/18/2025	10:54	LB135083
	Silver	-0.69			-10	10	03/18/2025	10:54	LB135083
ICSAB01	Arsenic	106	104	102	88.4	120	03/18/2025	11:04	LB135083
	Barium	452	537	84	437	637	03/18/2025	11:04	LB135083
	Cadmium	955	972	98	826	1120	03/18/2025	11:04	LB135083
	Chromium	555	542	102	460	624	03/18/2025	11:04	LB135083
	Lead	51.9	49.0	106	37	61	03/18/2025	11:04	LB135083
	Selenium	34.2	46.0	74	26	66	03/18/2025	11:04	LB135083
	Silver	211	201	105	170	232	03/18/2025	11:04	LB135083
ICSA01	Arsenic	4.77			-20	20	03/25/2025	10:27	LB135182
	Barium	-0.56	6.0	9	-94	106	03/25/2025	10:27	LB135182
	Cadmium	-3.57	1.0	357	-5	7	03/25/2025	10:27	LB135182
	Chromium	55.0	52.0	106	42	62	03/25/2025	10:27	LB135182
	Lead	4.19			-12	12	03/25/2025	10:27	LB135182
	Selenium	-9.76			-20	20	03/25/2025	10:27	LB135182
	Silver	-0.0040			-10	10	03/25/2025	10:27	LB135182
ICSAB01	Arsenic	112	104	108	88.4	120	03/25/2025	10:31	LB135182
	Barium	469	537	87	437	637	03/25/2025	10:31	LB135182
	Cadmium	998	972	103	826	1120	03/25/2025	10:31	LB135182
	Chromium	566	542	104	460	624	03/25/2025	10:31	LB135182
	Lead	52.0	49.0	106	37	61	03/25/2025	10:31	LB135182
	Selenium	37.0	46.0	80	26	66	03/25/2025	10:31	LB135182

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Weston Solutions **SDG No.:** Q1569
Contract: WEST04 **Lab Code:** CHEM **Case No.:** Q1569 **SAS No.:** Q1569
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Silver	203	201	101	170	232	03/25/2025	10:31	LB135182