

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB167214BS

Contract: WEST04

Lab Code: CHEM

Case No.: Q1569

SAS No.: Q1569

SDG NO.: Q1569

Lab Sample ID: PB167214BS

Date(s) Analyzed: 03/19/2025

03/19/2025

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP

ID: 0.32 (mm)

GC Column:(2): RTX-CLP2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
2,4,5-TP(Silvex)	1	9.15	9.10	9.20	5.10	6.1
	2	9.74	9.69	9.79	4.80	
2,4-D	1	8.28	8.23	8.33	4.70	6.6
	2	8.85	8.80	8.90	4.40	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

TAP-IDW-SOIL-031325-01MS

Contract: WEST04

Lab Code: CHEM **Case No.:** Q1569

SAS No.: Q1569 **SDG NO.:** Q1569

Lab Sample ID: Q1569-02MS

Date(s) Analyzed: 03/19/2025 03/19/2025

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP **ID:** 0.32 (mm)

GC Column:(2): RTX-CLP2 **ID:** 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
2,4-D	1	8.27	8.22	8.32	52.5	8.3
	2	8.83	8.78	8.88	48.3	
2,4,5-TP(Silvex)	1	9.13	9.08	9.18	52.3	9.6
	2	9.72	9.67	9.77	47.5	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

TAP-IDW-SOIL-031325-01MSD

Contract: WEST04

Lab Code: CHEM **Case No.:** Q1569

SAS No.: Q1569 **SDG NO.:** Q1569

Lab Sample ID: Q1569-02MSD

Date(s) Analyzed: 03/19/2025 03/19/2025

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP **ID:** 0.32 (mm)

GC Column:(2): RTX-CLP2 **ID:** 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
2,4-D	1	8.27	8.22	8.32	53.1	8.4
	2	8.83	8.78	8.88	48.8	
2,4,5-TP(Silvex)	1	9.13	9.08	9.18	53.1	9.3
	2	9.72	9.67	9.77	48.4	