

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1569 SAS No.: Q1569 SDG NO.: Q1569

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 02/21/2025 02/21/2025

Client Sample No.: CCAL01 Date Analyzed: 03/19/2025

Lab Sample No.: HSTDCCC750 Data File : PS029453.D Time Analyzed: 10:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.144	9.038	9.238	736.050	712.500	3.3
2,4-D	8.279	8.171	8.371	689.350	705.000	-2.2
2,4-DCAA	7.163	7.060	7.260	750.070	750.000	0.0

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1569 SAS No.: Q1569 SDG NO.: Q1569

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/21/2025 02/21/2025

Client Sample No.: CCAL01 Date Analyzed: 03/19/2025

Lab Sample No.: HSTDCCC750 Data File : PS029453.D Time Analyzed: 10:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.741	9.656	9.856	716.950	712.500	0.6
2,4-D	8.849	8.761	8.961	664.460	705.000	-5.8
2,4-DCAA	7.623	7.532	7.732	691.880	750.000	-7.7