

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM **Case No.:** Q1569 **SAS No.:** Q1569 **SDG NO.:** Q1569

GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 02/21/2025 02/21/2025

Client Sample No.: CCAL02 **Date Analyzed:** 03/19/2025

Lab Sample No.: HSTDCCC750 **Data File :** PS029459.D **Time Analyzed:** 19:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.133	9.038	9.238	752.960	712.500	5.7
2,4-D	8.270	8.171	8.371	716.180	705.000	1.6
2,4-DCAA	7.157	7.060	7.260	777.210	750.000	3.6

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1569 SAS No.: Q1569 SDG NO.: Q1569

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/21/2025 02/21/2025

Client Sample No.: CCAL02 Date Analyzed: 03/19/2025

Lab Sample No.: HSTDCCC750 Data File : PS029459.D Time Analyzed: 19:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.717	9.656	9.856	728.590	712.500	2.3
2,4-D	8.827	8.761	8.961	682.290	705.000	-3.2
2,4-DCAA	7.604	7.532	7.732	702.510	750.000	-6.3