

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1569 SAS No.: Q1569 SDG NO.: Q1569

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 02/21/2025 02/21/2025

Client Sample No.: CCAL03 Date Analyzed: 03/20/2025

Lab Sample No.: HSTDCCC750 Data File : PS029468.D Time Analyzed: 00:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.134	9.038	9.238	760.220	712.500	6.7
2,4-D	8.270	8.171	8.371	724.230	705.000	2.7
2,4-DCAA	7.157	7.060	7.260	787.230	750.000	5.0

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1569 SAS No.: Q1569 SDG NO.: Q1569

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/21/2025 02/21/2025

Client Sample No.: CCAL03 Date Analyzed: 03/20/2025

Lab Sample No.: HSTDCCC750 Data File : PS029468.D Time Analyzed: 00:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.714	9.656	9.856	744.080	712.500	4.4
2,4-D	8.824	8.761	8.961	689.410	705.000	-2.2
2,4-DCAA	7.603	7.532	7.732	725.400	750.000	-3.3