



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

**SDG No.:** Q1569

**Client:** Weston Solutions

**Analytical Method:** 8151A **Datafile :** PS029455.D

| Lab Sample ID | Parameter        | Spike | Result | Units | Rec | RPD | Qual | RPD  |     | Limits |     |
|---------------|------------------|-------|--------|-------|-----|-----|------|------|-----|--------|-----|
|               |                  |       |        |       |     |     |      | Qual | Low | High   | RPD |
| PB167214BS    | 2,4-D            | 5     | 4.70   | ug/L  | 94  |     |      |      | 45  | 152    |     |
|               | 2,4,5-TP(Silvex) | 5     | 5.10   | ug/L  | 102 |     |      |      | 51  | 134    |     |