

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1569 SAS No.: Q1569 SDG NO.: Q1569

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL01 Date Analyzed: 03/19/2025

Lab Sample No.: PSTDCCC050 Data File : PL094738.D Time Analyzed: 15:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.050	8.956	9.156	47.380	50.000	-5.2
Endrin	6.569	6.474	6.674	45.420	50.000	-9.2
gamma-BHC (Lindane)	4.324	4.227	4.427	50.000	50.000	0.0
Heptachlor	4.912	4.815	5.015	45.970	50.000	-8.1
Heptachlor epoxide	5.680	5.583	5.783	48.150	50.000	-3.7
Methoxychlor	7.496	7.400	7.600	46.070	50.000	-7.9
Tetrachloro-m-xylene	3.536	3.438	3.638	50.490	50.000	1.0

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1569 SAS No.: Q1569 SDG NO.: Q1569

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL01 Date Analyzed: 03/19/2025

Lab Sample No.: PSTDCCC050 Data File : PL094738.D Time Analyzed: 15:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.901	7.807	8.007	40.160	50.000	-19.7
Endrin	5.630	5.534	5.734	50.650	50.000	1.3
gamma-BHC (Lindane)	3.602	3.504	3.704	54.710	50.000	9.4
Heptachlor	3.939	3.842	4.042	49.600	50.000	-0.8
Heptachlor epoxide	4.721	4.625	4.825	52.310	50.000	4.6
Methoxychlor	6.603	6.507	6.707	44.080	50.000	-11.8
Tetrachloro-m-xylene	2.770	2.672	2.872	55.450	50.000	10.9