

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1569 SAS No.: Q1569 SDG NO.: Q1569

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 03/19/2025

Lab Sample No.: PSTDCCC050 Data File : PL094752.D Time Analyzed: 19:31

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.050	8.956	9.156	50.180	50.000	0.4
Endrin	6.571	6.474	6.674	41.560	50.000	-16.9
gamma-BHC (Lindane)	4.324	4.227	4.427	48.620	50.000	-2.8
Heptachlor	4.912	4.815	5.015	44.780	50.000	-10.4
Heptachlor epoxide	5.680	5.583	5.783	47.620	50.000	-4.8
Methoxychlor	7.497	7.400	7.600	44.090	50.000	-11.8
Tetrachloro-m-xylene	3.536	3.438	3.638	48.910	50.000	-2.2

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1569 SAS No.: Q1569 SDG NO.: Q1569

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 03/19/2025

Lab Sample No.: PSTDCCC050 Data File : PL094752.D Time Analyzed: 19:31

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	7.903	7.807	8.007	47.370	50.000	-5.3
Endrin	5.631	5.534	5.734	48.310	50.000	-3.4
gamma-BHC (Lindane)	3.602	3.504	3.704	51.930	50.000	3.9
Heptachlor	3.939	3.842	4.042	47.360	50.000	-5.3
Heptachlor epoxide	4.720	4.625	4.825	50.720	50.000	1.4
Methoxychlor	6.603	6.507	6.707	45.410	50.000	-9.2
Tetrachloro-m-xylene	2.770	2.672	2.872	52.230	50.000	4.5