

SOP ID : M1311-TCLP-15

SDG No : N/A

Weigh By : JP

Balance ID : WC SC-7

pH Meter ID : WC PH METER-1

Extraction By : JP

Filter By : JP

Pipette ID : WC

Tumbler ID : T-1

TCLP Filter ID : 115525

Start Prep Date : 03/14/2025 Time : 14:00

End Prep Date : 03/15/2025 Time : 08:30

Combination Ratio : 20

ZHE Cleaning Batch : ¹⁸N/A

Initial Room Temperature: 23 °C

Final Room Temperature: 22 °C

TCLP Technician Signature : ¹⁸

Supervisor By : ¹²

Standard Name	MLS USED	STD REF. # FROM LOG
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE U	Lot Number
TCLP-FLUID-1	N/A	WP110802
HCL-TCLP,1N	N/A	WP110803
HNO3-TCLP,1N	N/A	WP110804
pH Strips	W3172.	W1931,W1934,W3171,W3172
N/A	W1941,W1942	W3166,W1938,W1939,W1940,
1 Liter Amber	N/A	90424-08
120ml Plastic bottle	N/A	405130101
1:1 HNO3	N/A	MP84041

Extraction Conformance/Non-Conformance Comments:

Matrix spikes are added after filtration and before preservation. TUMBLER T-1 checked.30 rpm. Particle size reduction is not required. q1568-04 is used for MS-MSD.

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
03/17/25 10:00	¹⁸ / TCLP Room	¹⁸ / Ext
	Preparation Group	Analysis Group

¹² met n'g

TCLP EXTRACTION LOGPAGE

PB167133

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
PB167133TB	LEB133	06	N/A	2000	N/A	N/A	N/A	4.93	1.0	T-1
Q1507-04	50-MIDDLESEX-AVE	01	100.03	2000	N/A	N/A	N/A	4.5	1.5	T-1
Q1568-04	JC-03-03132025	02	100.02	2000	N/A	N/A	N/A	5.5	1.0	T-1
Q1569-02	TAP-IDW-SOIL-031325-01	03	100.01	2000	N/A	N/A	N/A	3.5	1.5	T-1
Q1569-05	TAP-IDW-SOIL-031325-02	04	100.02	2000	N/A	N/A	N/A	7.6	1.0	T-1
Q1574-02	WC1	05	100.03	2000	N/A	N/A	N/A	4.5	1.5	T-1

SampleID	ClientID	Sample Weight (g)	Filter Weight (g)	Filtrate (mL)	Filter + Solid (After 100°C)	% solids	% Dry Solids
PB167133TB	LEB133	N/A	N/A	N/A	N/A	N/A	N/A
Q1507-04	50-MIDDLESEX-AVE	N/A	N/A	N/A	N/A	100	N/A
Q1568-04	JC-03-03132025	N/A	N/A	N/A	N/A	100	N/A
Q1569-02	TAP-IDW-SOIL-031325-01	N/A	N/A	N/A	N/A	100	N/A
Q1569-05	TAP-IDW-SOIL-031325-02	N/A	N/A	N/A	N/A	100	N/A
Q1574-02	WC1	N/A	N/A	N/A	N/A	100	N/A

Hot Block ID : WC S-1 /WC S-2

 Thermometer ID : FLASHPOINT

SampleID	ClientID	Sample Weight (g)	Volume DI Water (mL)	PH after 5 min stir	PH after 10 min stir	Extraction Fluid 1 or 2	pH Extraction Fluid
PB167133TB	LEB133	N/A	N/A	N/A	N/A	#1	4.93
Q1507-04	50-MIDDLESEX-AVE	5.02	96.5	7.0	2.5	#1	4.93
Q1568-04	JC-03-03132025	5.01	96.5	7.0	2.5	#1	4.93
Q1569-02	TAP-IDW-SOIL-031325-01	5.03	96.5	6.8	2.0	#1	4.93
Q1569-05	TAP-IDW-SOIL-031325-02	5.04	96.5	10.0	4.0	#1	4.93
Q1574-02	WC1	5.03	96.5	6.8	2.0	#1	4.93