

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF031025

Review By	anahy	Review On	3/11/2025 9:11:47 AM
Supervise By	Jagrut	Supervise On	3/11/2025 10:18:40 AM
SubDirectory	BF031025	HP Acquire Method	BNA_F
		HP Processing Method	bf031025

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12653,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF141896.D	10 Mar 2025 10:31		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF141897.D	10 Mar 2025 11:01		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF141898.D	10 Mar 2025 11:30	Compound#9,32,54,65 removed from 5 ppm	RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF141899.D	10 Mar 2025 12:00		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF141900.D	10 Mar 2025 12:29	Compound #54 Kept on LR, Method is good for DOD . Method failed for compound #77.	RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF141901.D	10 Mar 2025 12:58		RC/JU	Ok
7	SSTDICC050	SSTDICC050	BF141902.D	10 Mar 2025 13:28	Not used	RC/JU	Not Ok
8	SSTDICC060	SSTDICC060	BF141903.D	10 Mar 2025 13:57	Compound#69 Removed from 60 ppm	RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF141904.D	10 Mar 2025 14:27	Compound#9,69 Removed from 80 ppm	RC/JU	Ok,M
10	SSTDICC050	SSTDICC050	BF141905.D	10 Mar 2025 15:20		RC/JU	Ok
11	SSTDICV040	ICVBF031025	BF141906.D	10 Mar 2025 15:53		RC/JU	Ok
12	PB167012BL	PB167012BL	BF141907.D	10 Mar 2025 17:09		RC/JU	Ok
13	SP6752	SP6752	BF141908.D	10 Mar 2025 17:39	8270 SPIKE-SP6752	RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF031825

Review By	anahy	Review On	3/19/2025 8:59:28 AM
Supervise By	Jagrut	Supervise On	3/20/2025 5:01:44 PM
SubDirectory	BF031825	HP Acquire Method	BNA_F
		HP Processing Method	bf031025

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12655,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF141985.D	18 Mar 2025 09:31		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF141986.D	18 Mar 2025 10:00		RC/JU	Ok
3	PB167168BL	PB167168BL	BF141987.D	18 Mar 2025 10:30		RC/JU	Ok
4	PB167168BS	PB167168BS	BF141988.D	18 Mar 2025 10:59		RC/JU	Ok,M
5	PB167133TB	PB167133TB	BF141989.D	18 Mar 2025 11:28		RC/JU	Ok
6	Q1569-02	TAP-IDW-SOIL-031325	BF141990.D	18 Mar 2025 12:03		RC/JU	Ok
7	Q1569-05	TAP-IDW-SOIL-031325	BF141991.D	18 Mar 2025 12:32		RC/JU	Ok
8	Q1568-04MS	JC-03-03132025MS	BF141992.D	18 Mar 2025 13:01		RC/JU	Ok,M
9	Q1568-04MSD	JC-03-03132025MSD	BF141993.D	18 Mar 2025 13:31		RC/JU	Ok,M
10	SSTDCCC040	SSTDCCC040EC	BF141994.D	18 Mar 2025 14:30		RC/JU	Ok

M : Manual Integration