

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1569

Client: Weston Solutions

Analytical Method: 8270E

DataFile: BF141988.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB167168BS	Pyridine	50	40.0	ug/L	80				10	153	
	1,4-Dichlorobenzene	50	45.5	ug/L	91				29	112	
	2-Methylphenol	50	45.8	ug/L	92				30	117	
	3+4-Methylphenols	50	44.7	ug/L	89				29	110	
	Hexachloroethane	50	44.6	ug/L	89				21	115	
	Nitrobenzene	50	44.6	ug/L	89				45	121	
	Hexachlorobutadiene	50	46.1	ug/L	92				22	124	
	2,4,6-Trichlorophenol	50	45.5	ug/L	91				50	125	
	2,4,5-Trichlorophenol	50	46.9	ug/L	94				53	123	
	2,4-Dinitrotoluene	50	47.6	ug/L	95				57	128	
	Hexachlorobenzene	50	48.3	ug/L	97				53	125	
	Pentachlorophenol	100	93.4	ug/L	93				35	138	