

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: WEST04

Lab Code: CHEM Case No.: Q1569 SAS No.: Q1569 SDG No.: Q1569

Instrument ID: MSVOA\_X Calibration Date/Time: 03/17/2025 11:12

Lab File ID: VX045297.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

| COMPOUND              | RRF   | RRF050 | MIN<br>RRF | %D    | MAX%D |
|-----------------------|-------|--------|------------|-------|-------|
| Vinyl Chloride        | 0.764 | 0.641  |            | -16.1 | 20    |
| 1,1-Dichloroethene    | 0.614 | 0.587  |            | -4.4  | 20    |
| 2-Butanone            | 0.555 | 0.563  |            | 1.44  | 20    |
| Carbon Tetrachloride  | 0.465 | 0.487  |            | 4.73  | 20    |
| Chloroform            | 1.239 | 1.179  |            | -4.84 | 20    |
| Benzene               | 1.448 | 1.393  |            | -3.8  | 20    |
| 1,2-Dichloroethane    | 0.521 | 0.523  |            | 0.38  | 20    |
| Trichloroethene       | 0.340 | 0.326  |            | -4.12 | 20    |
| Tetrachloroethene     | 0.320 | 0.314  |            | -1.88 | 20    |
| Chlorobenzene         | 1.058 | 1.062  | 0.3        | 0.38  | 20    |
| 1,2-Dichloroethane-d4 | 0.795 | 0.805  |            | 1.26  | 20    |
| Dibromofluoromethane  | 0.334 | 0.347  |            | 3.89  | 20    |
| Toluene-d8            | 1.212 | 1.199  |            | -1.07 | 20    |
| 4-Bromofluorobenzene  | 0.402 | 0.434  |            | 7.96  | 20    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.