

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: WEST04

Lab Code: CHEM Case No.: Q1569 SAS No.: Q1569 SDG No.: Q1569

Instrument ID: MSVOA_X Calibration Date/Time: 03/17/2025 16:58

Lab File ID: VX045310.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.764	0.622		-18.59	50
1,1-Dichloroethene	0.614	0.587		-4.4	50
2-Butanone	0.555	0.578		4.14	50
Carbon Tetrachloride	0.465	0.476		2.37	50
Chloroform	1.239	1.242		0.24	50
Benzene	1.448	1.431		-1.17	50
1,2-Dichloroethane	0.521	0.550		5.57	50
Trichloroethene	0.340	0.331		-2.65	50
Tetrachloroethene	0.320	0.303		-5.31	50
Chlorobenzene	1.058	1.070	0.3	1.13	50
1,2-Dichloroethane-d4	0.795	0.828		4.15	50
Dibromofluoromethane	0.334	0.350		4.79	50
Toluene-d8	1.212	1.232		1.65	50
4-Bromofluorobenzene	0.402	0.448		11.44	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.