

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1569
Client: Weston Solutions
Analytical Method: SW8260D **Datafile :** VX045301.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0317WBS01	Vinyl chloride	20	17.1	ug/L	86			58	137	
	1,1-Dichloroethene	20	20.1	ug/L	101			71	131	
	2-Butanone	100	110	ug/L	110			56	143	
	Carbon Tetrachloride	20	20.0	ug/L	100			72	136	
	Chloroform	20	21.1	ug/L	106			79	124	
	Benzene	20	19.8	ug/L	99			79	120	
	1,2-Dichloroethane	20	21.6	ug/L	108			73	128	
	Trichloroethene	20	19.6	ug/L	98			79	123	
	Tetrachloroethene	20	19.6	ug/L	98			74	129	
	Chlorobenzene	20	20.8	ug/L	104			82	118	