

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1569

Client: Weston Solutions

Analytical Method: SW8260D

Datafile : VX045302.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0317WBSD01	Vinyl chloride	20	15.8	ug/L	79	8		58	137	20
	1,1-Dichloroethene	20	18.9	ug/L	95	6		71	131	20
	2-Butanone	100	110	ug/L	110	0		56	143	20
	Carbon Tetrachloride	20	19.8	ug/L	99	1		72	136	20
	Chloroform	20	20.7	ug/L	104	2		79	124	20
	Benzene	20	20.2	ug/L	101	2		79	120	20
	1,2-Dichloroethane	20	22.6	ug/L	113	5		73	128	20
	Trichloroethene	20	19.5	ug/L	98	0		79	123	20
	Tetrachloroethene	20	19.6	ug/L	98	0		74	129	20
	Chlorobenzene	20	20.7							