

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WEST04
 Lab Code: CHEM Case No.: Q1569 SAS No.: Q1569 SDG NO.: Q1569
 Lab File ID: VX045297.D Date Analyzed: 03/17/2025
 Instrument ID: MSVOA_X Time Analyzed: 11:12
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	113376	5.54	196865	6.75	167918	10.05
UPPER LIMIT	226752	6.038	393730	7.251	335836	10.549
LOWER LIMIT	56688	5.038	98432.5	6.251	83959	9.549
EPA SAMPLE NO.						
TAP-IDW-SOIL-031325-01	76405	5.55	150178	6.76	135949	10.06
TAP-IDW-SOIL-031325-02	71679	5.54	146807	6.76	133342	10.05
VX0317WBL01	73730	5.55	144901	6.76	130224	10.05
VX0317WBS01	101138	5.54	185864	6.76	165274	10.05
VX0317WBSD01	97711	5.54	175105	6.76	156081	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.