

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WEST04
 Lab Code: CHEM Case No.: Q1569 SAS No.: Q1569 SDG NO.: Q1569
 Lab File ID: VX045297.D Date Analyzed: 03/17/2025
 Instrument ID: MSVOA_X Time Analyzed: 11:12
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	79039	12.018				
UPPER LIMIT	158078	12.518				
LOWER LIMIT	39519.5	11.518				
EPA SAMPLE NO.						
TAP-IDW-SOIL-031325-01	53669	12.02				
TAP-IDW-SOIL-031325-02	52735	12.02				
VX0317WBL01	50695	12.02				
VX0317WBS01	73450	12.02				
VX0317WBSD01	68359	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.