

### CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1572 SAS No.: Q1572 SDG NO.: Q1572

Continuing Calib Date: 03/15/2025 Initial Calibration Date(s): 02/20/2025 02/21/2025

Continuing Calib Time: 04:46 Initial Calibration Time(s): 16:46 01:02

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1242-1 (1)   | 4.79       | 4.79      | 4.69      | 4.89 | 0.00       |
| Aroclor-1242-2 (2)   | 4.81       | 4.81      | 4.71      | 4.91 | 0.01       |
| Aroclor-1242-3 (3)   | 4.86       | 4.87      | 4.77      | 4.97 | 0.01       |
| Aroclor-1242-4 (4)   | 4.98       | 4.99      | 4.89      | 5.09 | 0.01       |
| Aroclor-1242-5 (5)   | 5.64       | 5.64      | 5.54      | 5.74 | 0.00       |
| Tetrachloro-m-xylene | 3.69       | 3.70      | 3.60      | 3.80 | 0.01       |
| Decachlorobiphenyl   | 8.74       | 8.76      | 8.66      | 8.86 | 0.02       |

### CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1572 SAS No.: Q1572 SDG NO.: Q1572

Continuing Calib Date: 03/15/2025 Initial Calibration Date(s): 02/20/2025 02/21/2025

Continuing Calib Time: 04:46 Initial Calibration Time(s): 16:46 01:02

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1242-1       | (1) | 4.77       | 4.78      | 4.68      | 4.88 | 0.01       |
| Aroclor-1242-2       | (2) | 4.79       | 4.80      | 4.70      | 4.90 | 0.01       |
| Aroclor-1242-3       | (3) | 4.97       | 4.97      | 4.87      | 5.07 | 0.00       |
| Aroclor-1242-4       | (4) | 5.05       | 5.06      | 4.96      | 5.16 | 0.01       |
| Aroclor-1242-5       | (5) | 5.57       | 5.58      | 5.48      | 5.68 | 0.01       |
| Tetrachloro-m-xylene |     | 3.69       | 3.69      | 3.59      | 3.79 | 0.00       |
| Decachlorobiphenyl   |     | 8.70       | 8.71      | 8.61      | 8.81 | 0.02       |