

### CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1572 SAS No.: Q1572 SDG NO.: Q1572

Continuing Calib Date: 03/17/2025 Initial Calibration Date(s): 03/11/2025 03/11/2025

Continuing Calib Time: 09:41 Initial Calibration Time(s): 15:13 23:38

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|------------|-----------|-----------|-------|------------|
|                      |            |           | FROM      | TO    |            |
| Aroclor-1248-1 (1)   | 5.67       | 5.68      | 5.58      | 5.78  | 0.01       |
| Aroclor-1248-2 (2)   | 5.95       | 5.95      | 5.85      | 6.05  | 0.00       |
| Aroclor-1248-3 (3)   | 6.15       | 6.15      | 6.05      | 6.25  | 0.00       |
| Aroclor-1248-4 (4)   | 6.55       | 6.55      | 6.45      | 6.65  | 0.00       |
| Aroclor-1248-5 (5)   | 6.59       | 6.59      | 6.49      | 6.69  | 0.00       |
| Tetrachloro-m-xylene | 4.52       | 4.52      | 4.42      | 4.62  | 0.00       |
| Decachlorobiphenyl   | 10.24      | 10.25     | 10.15     | 10.35 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1572 SAS No.: Q1572 SDG NO.: Q1572

Continuing Calib Date: 03/17/2025 Initial Calibration Date(s): 03/11/2025 03/11/2025

Continuing Calib Time: 09:41 Initial Calibration Time(s): 15:13 23:38

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1248-1       | (1) | 4.91       | 4.92      | 4.82      | 5.02 | 0.01       |
| Aroclor-1248-2       | (2) | 5.15       | 5.15      | 5.05      | 5.25 | 0.00       |
| Aroclor-1248-3       | (3) | 5.19       | 5.20      | 5.10      | 5.30 | 0.01       |
| Aroclor-1248-4       | (4) | 5.37       | 5.37      | 5.27      | 5.47 | 0.01       |
| Aroclor-1248-5       | (5) | 5.76       | 5.76      | 5.66      | 5.86 | 0.00       |
| Tetrachloro-m-xylene |     | 3.83       | 3.83      | 3.73      | 3.93 | 0.00       |
| Decachlorobiphenyl   |     | 8.87       | 8.88      | 8.78      | 8.98 | 0.01       |