

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1572 SAS No.: Q1572 SDG NO.: Q1572

Continuing Calib Date: 03/17/2025 Initial Calibration Date(s): 03/11/2025 03/11/2025

Continuing Calib Time: 14:49 Initial Calibration Time(s): 15:13 23:38

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	6.53	6.53	6.43	6.63	0.00
Aroclor-1254-2	(2)	6.74	6.75	6.65	6.85	0.01
Aroclor-1254-3	(3)	7.11	7.11	7.01	7.21	0.00
Aroclor-1254-4	(4)	7.39	7.39	7.29	7.49	0.00
Aroclor-1254-5	(5)	7.81	7.81	7.71	7.91	0.00
Tetrachloro-m-xylene		4.53	4.53	4.43	4.63	0.00
Decachlorobiphenyl		10.25	10.25	10.15	10.35	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1572 SAS No.: Q1572 SDG NO.: Q1572

Continuing Calib Date: 03/17/2025 Initial Calibration Date(s): 03/11/2025 03/11/2025

Continuing Calib Time: 14:49 Initial Calibration Time(s): 15:13 23:38

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	5.72	5.72	5.62	5.82	0.00
Aroclor-1254-2	(2)	5.87	5.87	5.77	5.97	0.00
Aroclor-1254-3	(3)	6.27	6.27	6.17	6.37	0.00
Aroclor-1254-4	(4)	6.50	6.50	6.40	6.60	0.00
Aroclor-1254-5	(5)	6.92	6.92	6.82	7.02	0.00
Tetrachloro-m-xylene		3.83	3.83	3.73	3.93	0.00
Decachlorobiphenyl		8.87	8.88	8.78	8.98	0.01