

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1572 SAS No.: Q1572 SDG NO.: Q1572

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL01 Date Analyzed: 03/14/2025

Lab Sample No.: AR1660CCC500 Data File : PO109890.D Time Analyzed: 14:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.786	4.691	4.891	508.050	500.000	1.6
Aroclor-1016-2	4.805	4.711	4.911	517.850	500.000	3.6
Aroclor-1016-3	4.862	4.767	4.967	514.830	500.000	3.0
Aroclor-1016-4	4.982	4.888	5.088	514.530	500.000	2.9
Aroclor-1016-5	5.240	5.145	5.345	562.950	500.000	12.6
Aroclor-1260-1	6.280	6.188	6.388	481.900	500.000	-3.6
Aroclor-1260-2	6.469	6.377	6.577	471.940	500.000	-5.6
Aroclor-1260-3	6.837	6.746	6.946	469.390	500.000	-6.1
Aroclor-1260-4	7.097	7.005	7.205	468.700	500.000	-6.3
Aroclor-1260-5	7.339	7.247	7.447	459.080	500.000	-8.2
Decachlorobiphenyl	8.744	8.657	8.857	40.060	50.000	-19.9
Tetrachloro-m-xylene	3.694	3.598	3.798	49.190	50.000	-1.6

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1572 SAS No.: Q1572 SDG NO.: Q1572

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL01 Date Analyzed: 03/14/2025

Lab Sample No.: AR1660CCC500 Data File : PO109890.D Time Analyzed: 14:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.772	4.678	4.878	544.710	500.000	8.9
Aroclor-1016-2	4.792	4.697	4.897	550.290	500.000	10.1
Aroclor-1016-3	4.967	4.872	5.072	547.520	500.000	9.5
Aroclor-1016-4	5.009	4.914	5.114	451.940	500.000	-9.6
Aroclor-1016-5	5.222	5.127	5.327	597.560	500.000	19.5
Aroclor-1260-1	6.254	6.161	6.361	512.390	500.000	2.5
Aroclor-1260-2	6.441	6.348	6.548	507.480	500.000	1.5
Aroclor-1260-3	6.595	6.502	6.702	488.590	500.000	-2.3
Aroclor-1260-4	7.066	6.974	7.174	475.030	500.000	-5.0
Aroclor-1260-5	7.306	7.214	7.414	480.380	500.000	-3.9
Decachlorobiphenyl	8.695	8.607	8.807	39.610	50.000	-20.8
Tetrachloro-m-xylene	3.691	3.594	3.794	51.910	50.000	3.8