

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1572 SAS No.: Q1572 SDG NO.: Q1572

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL02 Date Analyzed: 03/14/2025

Lab Sample No.: AR1660CCC500 Data File : PO109904.D Time Analyzed: 21:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.786	4.691	4.891	535.740	500.000	7.1
Aroclor-1016-2	4.805	4.711	4.911	547.300	500.000	9.5
Aroclor-1016-3	4.861	4.767	4.967	537.930	500.000	7.6
Aroclor-1016-4	4.982	4.888	5.088	536.100	500.000	7.2
Aroclor-1016-5	5.239	5.145	5.345	524.580	500.000	4.9
Aroclor-1260-1	6.279	6.188	6.388	509.750	500.000	2.0
Aroclor-1260-2	6.468	6.377	6.577	495.260	500.000	-0.9
Aroclor-1260-3	6.836	6.746	6.946	469.260	500.000	-6.1
Aroclor-1260-4	7.096	7.005	7.205	457.140	500.000	-8.6
Aroclor-1260-5	7.338	7.247	7.447	454.720	500.000	-9.1
Decachlorobiphenyl	8.743	8.657	8.857	39.340	50.000	-21.3
Tetrachloro-m-xylene	3.694	3.598	3.798	50.970	50.000	1.9

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1572 SAS No.: Q1572 SDG NO.: Q1572

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL02 Date Analyzed: 03/14/2025

Lab Sample No.: AR1660CCC500 Data File : PO109904.D Time Analyzed: 21:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.772	4.678	4.878	559.670	500.000	11.9
Aroclor-1016-2	4.792	4.697	4.897	576.210	500.000	15.2
Aroclor-1016-3	4.967	4.872	5.072	566.880	500.000	13.4
Aroclor-1016-4	5.009	4.914	5.114	533.730	500.000	6.7
Aroclor-1016-5	5.222	5.127	5.327	553.280	500.000	10.7
Aroclor-1260-1	6.254	6.161	6.361	510.600	500.000	2.1
Aroclor-1260-2	6.442	6.348	6.548	513.060	500.000	2.6
Aroclor-1260-3	6.594	6.502	6.702	501.050	500.000	0.2
Aroclor-1260-4	7.065	6.974	7.174	485.160	500.000	-3.0
Aroclor-1260-5	7.306	7.214	7.414	495.830	500.000	-0.8
Decachlorobiphenyl	8.696	8.607	8.807	41.110	50.000	-17.8
Tetrachloro-m-xylene	3.690	3.594	3.794	51.780	50.000	3.6