

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1572 SAS No.: Q1572 SDG NO.: Q1572

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL03 Date Analyzed: 03/17/2025

Lab Sample No.: AR1660CCC500 Data File : PO109923.D Time Analyzed: 09:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.787	4.691	4.891	520.420	500.000	4.1
Aroclor-1016-2	4.806	4.711	4.911	532.150	500.000	6.4
Aroclor-1016-3	4.862	4.767	4.967	529.790	500.000	6.0
Aroclor-1016-4	4.983	4.888	5.088	529.650	500.000	5.9
Aroclor-1016-5	5.241	5.145	5.345	513.580	500.000	2.7
Aroclor-1260-1	6.281	6.188	6.388	491.640	500.000	-1.7
Aroclor-1260-2	6.470	6.377	6.577	488.670	500.000	-2.3
Aroclor-1260-3	6.838	6.746	6.946	474.600	500.000	-5.1
Aroclor-1260-4	7.098	7.005	7.205	472.840	500.000	-5.4
Aroclor-1260-5	7.340	7.247	7.447	469.530	500.000	-6.1
Decachlorobiphenyl	8.745	8.657	8.857	40.020	50.000	-20.0
Tetrachloro-m-xylene	3.694	3.598	3.798	52.550	50.000	5.1

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1572 SAS No.: Q1572 SDG NO.: Q1572

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL03 Date Analyzed: 03/17/2025

Lab Sample No.: AR1660CCC500 Data File : PO109923.D Time Analyzed: 09:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.772	4.678	4.878	539.290	500.000	7.9
Aroclor-1016-2	4.792	4.697	4.897	554.740	500.000	10.9
Aroclor-1016-3	4.967	4.872	5.072	548.600	500.000	9.7
Aroclor-1016-4	5.009	4.914	5.114	529.410	500.000	5.9
Aroclor-1016-5	5.222	5.127	5.327	523.530	500.000	4.7
Aroclor-1260-1	6.254	6.161	6.361	502.590	500.000	0.5
Aroclor-1260-2	6.442	6.348	6.548	501.730	500.000	0.3
Aroclor-1260-3	6.595	6.502	6.702	496.860	500.000	-0.6
Aroclor-1260-4	7.066	6.974	7.174	484.000	500.000	-3.2
Aroclor-1260-5	7.307	7.214	7.414	488.740	500.000	-2.3
Decachlorobiphenyl	8.696	8.607	8.807	39.180	50.000	-21.6
Tetrachloro-m-xylene	3.691	3.594	3.794	52.290	50.000	4.6