

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1572 SAS No.: Q1572 SDG NO.: Q1572

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL05 Date Analyzed: 03/18/2025

Lab Sample No.: AR1660CCC500 Data File : PO109963.D Time Analyzed: 08:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.785	4.691	4.891	523.130	500.000	4.6
Aroclor-1016-2	4.804	4.711	4.911	523.500	500.000	4.7
Aroclor-1016-3	4.860	4.767	4.967	527.350	500.000	5.5
Aroclor-1016-4	4.981	4.888	5.088	526.280	500.000	5.3
Aroclor-1016-5	5.238	5.145	5.345	574.940	500.000	15.0
Aroclor-1260-1	6.280	6.188	6.388	502.140	500.000	0.4
Aroclor-1260-2	6.469	6.377	6.577	491.200	500.000	-1.8
Aroclor-1260-3	6.837	6.746	6.946	471.160	500.000	-5.8
Aroclor-1260-4	7.096	7.005	7.205	467.620	500.000	-6.5
Aroclor-1260-5	7.339	7.247	7.447	459.320	500.000	-8.1
Decachlorobiphenyl	8.746	8.657	8.857	56.250	50.000	12.5
Tetrachloro-m-xylene	3.692	3.598	3.798	50.350	50.000	0.7

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1572 SAS No.: Q1572 SDG NO.: Q1572

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL05 Date Analyzed: 03/18/2025

Lab Sample No.: AR1660CCC500 Data File : PO109963.D Time Analyzed: 08:06

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.770	4.678	4.878	517.580	500.000	3.5
Aroclor-1016-2	4.789	4.697	4.897	534.310	500.000	6.9
Aroclor-1016-3	4.965	4.872	5.072	530.300	500.000	6.1
Aroclor-1016-4	5.007	4.914	5.114	451.310	500.000	-9.7
Aroclor-1016-5	5.220	5.127	5.327	569.280	500.000	13.9
Aroclor-1260-1	6.252	6.161	6.361	496.540	500.000	-0.7
Aroclor-1260-2	6.440	6.348	6.548	496.670	500.000	-0.7
Aroclor-1260-3	6.593	6.502	6.702	478.930	500.000	-4.2
Aroclor-1260-4	7.065	6.974	7.174	459.430	500.000	-8.1
Aroclor-1260-5	7.305	7.214	7.414	456.320	500.000	-8.7
Decachlorobiphenyl	8.695	8.607	8.807	47.180	50.000	-5.6
Tetrachloro-m-xylene	3.689	3.594	3.794	49.850	50.000	-0.3