

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1572 SAS No.: Q1572 SDG NO.: Q1572

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL07 Date Analyzed: 03/14/2025

Lab Sample No.: AR1660CCC500 Data File : PP070543.D Time Analyzed: 21:15

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.676	5.577	5.777	443.660	500.000	-11.3
Aroclor-1016-2	5.698	5.599	5.799	480.930	500.000	-3.8
Aroclor-1016-3	5.760	5.661	5.861	466.910	500.000	-6.6
Aroclor-1016-4	5.858	5.758	5.958	456.460	500.000	-8.7
Aroclor-1016-5	6.150	6.051	6.251	453.110	500.000	-9.4
Aroclor-1260-1	7.269	7.171	7.371	470.290	500.000	-5.9
Aroclor-1260-2	7.523	7.425	7.625	461.520	500.000	-7.7
Aroclor-1260-3	7.881	7.783	7.983	466.760	500.000	-6.6
Aroclor-1260-4	8.105	8.008	8.208	454.640	500.000	-9.1
Aroclor-1260-5	8.425	8.328	8.528	467.430	500.000	-6.5
Decachlorobiphenyl	10.243	10.152	10.352	46.560	50.000	-6.9
Tetrachloro-m-xylene	4.525	4.424	4.624	46.830	50.000	-6.3

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1572 SAS No.: Q1572 SDG NO.: Q1572

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL07 Date Analyzed: 03/14/2025

Lab Sample No.: AR1660CCC500 Data File : PP070543.D Time Analyzed: 21:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.912	4.817	5.017	467.760	500.000	-6.4
Aroclor-1016-2	4.932	4.836	5.036	453.880	500.000	-9.2
Aroclor-1016-3	5.108	5.013	5.213	459.150	500.000	-8.2
Aroclor-1016-4	5.150	5.054	5.254	462.340	500.000	-7.5
Aroclor-1016-5	5.365	5.270	5.470	472.460	500.000	-5.5
Aroclor-1260-1	6.401	6.307	6.507	458.750	500.000	-8.3
Aroclor-1260-2	6.589	6.495	6.695	458.120	500.000	-8.4
Aroclor-1260-3	6.743	6.649	6.849	450.000	500.000	-10.0
Aroclor-1260-4	7.215	7.121	7.321	441.280	500.000	-11.7
Aroclor-1260-5	7.456	7.362	7.562	440.290	500.000	-11.9
Decachlorobiphenyl	8.871	8.781	8.981	41.930	50.000	-16.1
Tetrachloro-m-xylene	3.826	3.729	3.929	48.070	50.000	-3.9