

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1572 SAS No.: Q1572 SDG NO.: Q1572

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL10 Date Analyzed: 03/17/2025

Lab Sample No.: AR1660CCC500 Data File : PP070574.D Time Analyzed: 09:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.676	5.577	5.777	467.050	500.000	-6.6
Aroclor-1016-2	5.697	5.599	5.799	503.160	500.000	0.6
Aroclor-1016-3	5.759	5.661	5.861	487.760	500.000	-2.4
Aroclor-1016-4	5.857	5.758	5.958	471.550	500.000	-5.7
Aroclor-1016-5	6.150	6.051	6.251	491.600	500.000	-1.7
Aroclor-1260-1	7.268	7.171	7.371	499.390	500.000	-0.1
Aroclor-1260-2	7.522	7.425	7.625	498.690	500.000	-0.3
Aroclor-1260-3	7.881	7.783	7.983	504.590	500.000	0.9
Aroclor-1260-4	8.105	8.008	8.208	489.180	500.000	-2.2
Aroclor-1260-5	8.425	8.328	8.528	502.500	500.000	0.5
Decachlorobiphenyl	10.245	10.152	10.352	50.640	50.000	1.3
Tetrachloro-m-xylene	4.524	4.424	4.624	49.130	50.000	-1.7

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1572 SAS No.: Q1572 SDG NO.: Q1572

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL10 Date Analyzed: 03/17/2025

Lab Sample No.: AR1660CCC500 Data File : PP070574.D Time Analyzed: 09:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.913	4.817	5.017	511.980	500.000	2.4
Aroclor-1016-2	4.932	4.836	5.036	496.640	500.000	-0.7
Aroclor-1016-3	5.109	5.013	5.213	504.730	500.000	0.9
Aroclor-1016-4	5.150	5.054	5.254	515.160	500.000	3.0
Aroclor-1016-5	5.365	5.270	5.470	500.200	500.000	0.0
Aroclor-1260-1	6.402	6.307	6.507	495.840	500.000	-0.8
Aroclor-1260-2	6.590	6.495	6.695	490.030	500.000	-2.0
Aroclor-1260-3	6.744	6.649	6.849	484.360	500.000	-3.1
Aroclor-1260-4	7.216	7.121	7.321	470.310	500.000	-5.9
Aroclor-1260-5	7.457	7.362	7.562	472.850	500.000	-5.4
Decachlorobiphenyl	8.873	8.781	8.981	46.620	50.000	-6.8
Tetrachloro-m-xylene	3.826	3.729	3.929	51.220	50.000	2.4