

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0022025\
 Data File : P0109457.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Feb 2025 02:15
 Operator : YP/AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO022025AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 21 03:06:52 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 21 03:04:52 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.696	3.693	499.8E6	265.5E6	50.778	50.913
2) SA Decachlor...	8.755	8.705	431.6E6	154.0E6	50.609	50.133
Target Compounds						
26) L6 AR-1254-1	5.600	5.579	259.2E6	131.6E6	503.961	504.534
27) L6 AR-1254-2	5.749	5.727	227.8E6	116.8E6	504.590	506.344
28) L6 AR-1254-3	6.155	6.129	367.1E6	181.9E6	505.885	503.874
29) L6 AR-1254-4	6.385	6.356	217.7E6	101.2E6	507.824	506.393
30) L6 AR-1254-5	6.805	6.775	319.4E6	149.4E6	505.658	504.131



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO022025\
Data File : PO109457.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Feb 2025 02:15
Operator : YP/AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO022025AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 21 03:06:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO022025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 21 03:04:52 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.696 min
Delta R.T.: 0.000 min
Response: 499836032
Conc: 50.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO022025AR1254




#1 Tetrachloro-m-xylene

R.T.: 3.693 min
Delta R.T.: 0.000 min
Response: 265489951
Conc: 50.91 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.755 min
Delta R.T.: 0.000 min
Response: 431607149
Conc: 50.61 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.705 min
Delta R.T.: 0.000 min
Response: 154033700
Conc: 50.13 ng/ml



#26 AR-1254-1


R.T.: 5.600 min
Delta R.T.: 0.000 min
Response: 259182509
Conc: 503.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO022025AR1254




#26 AR-1254-1

R.T.: 5.579 min
Delta R.T.: 0.000 min
Response: 131640242
Conc: 504.53 ng/ml




#27 AR-1254-2

R.T.: 5.749 min
Delta R.T.: 0.000 min
Response: 227813435
Conc: 504.59 ng/ml



#27 AR-1254-2


R.T.: 5.727 min
Delta R.T.: 0.000 min
Response: 116803575
Conc: 506.34 ng/ml



#28 AR-1254-3


R.T.: 6.155 min
Delta R.T.: 0.000 min
Response: 367084723
Conc: 505.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO022025AR1254




#28 AR-1254-3

R.T.: 6.129 min
Delta R.T.: 0.000 min
Response: 181873790
Conc: 503.87 ng/ml




#29 AR-1254-4

R.T.: 6.385 min
Delta R.T.: 0.000 min
Response: 217729027
Conc: 507.82 ng/ml



#29 AR-1254-4


R.T.: 6.356 min
Delta R.T.: 0.000 min
Response: 101150355
Conc: 506.39 ng/ml



#30 AR-1254-5

R.T.: 6.805 min
Delta R.T.: 0.000 min
Response: 319385335
Conc: 505.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO022025AR1254



#30 AR-1254-5

R.T.: 6.775 min
Delta R.T.: 0.000 min
Response: 149412622
Conc: 504.13 ng/ml