

### CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1573 SAS No.: Q1573 SDG NO.: Q1573

Continuing Calib Date: 03/17/2025 Initial Calibration Date(s): 03/11/2025 03/11/2025

Continuing Calib Time: 09:57 Initial Calibration Time(s): 15:13 23:38

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1254-1       | (1) | 6.53       | 6.53      | 6.43      | 6.63  | 0.00       |
| Aroclor-1254-2       | (2) | 6.74       | 6.75      | 6.65      | 6.85  | 0.01       |
| Aroclor-1254-3       | (3) | 7.11       | 7.11      | 7.01      | 7.21  | 0.00       |
| Aroclor-1254-4       | (4) | 7.39       | 7.39      | 7.29      | 7.49  | 0.00       |
| Aroclor-1254-5       | (5) | 7.80       | 7.81      | 7.71      | 7.91  | 0.01       |
| Tetrachloro-m-xylene |     | 4.52       | 4.53      | 4.43      | 4.63  | 0.01       |
| Decachlorobiphenyl   |     | 10.25      | 10.25     | 10.15     | 10.35 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1573 SAS No.: Q1573 SDG NO.: Q1573

Continuing Calib Date: 03/17/2025 Initial Calibration Date(s): 03/11/2025 03/11/2025

Continuing Calib Time: 09:57 Initial Calibration Time(s): 15:13 23:38

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1254-1 (1)   | 5.72       | 5.72      | 5.62      | 5.82 | 0.00       |
| Aroclor-1254-2 (2)   | 5.87       | 5.87      | 5.77      | 5.97 | 0.00       |
| Aroclor-1254-3 (3)   | 6.27       | 6.27      | 6.17      | 6.37 | 0.00       |
| Aroclor-1254-4 (4)   | 6.50       | 6.50      | 6.40      | 6.60 | 0.00       |
| Aroclor-1254-5 (5)   | 6.92       | 6.92      | 6.82      | 7.02 | 0.00       |
| Tetrachloro-m-xylene | 3.83       | 3.83      | 3.73      | 3.93 | 0.00       |
| Decachlorobiphenyl   | 8.87       | 8.88      | 8.78      | 8.98 | 0.01       |