

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1573 SAS No.: Q1573 SDG NO.: Q1573

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL01 Date Analyzed: 03/17/2025

Lab Sample No.: AR1660CCC500 Data File : PP070589.D Time Analyzed: 14:00

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.674	5.577	5.777	464.700	500.000	-7.1
Aroclor-1016-2	5.696	5.599	5.799	510.500	500.000	2.1
Aroclor-1016-3	5.759	5.661	5.861	489.260	500.000	-2.1
Aroclor-1016-4	5.856	5.758	5.958	488.460	500.000	-2.3
Aroclor-1016-5	6.149	6.051	6.251	489.200	500.000	-2.2
Aroclor-1260-1	7.268	7.171	7.371	520.980	500.000	4.2
Aroclor-1260-2	7.521	7.425	7.625	510.080	500.000	2.0
Aroclor-1260-3	7.880	7.783	7.983	512.770	500.000	2.6
Aroclor-1260-4	8.104	8.008	8.208	494.890	500.000	-1.0
Aroclor-1260-5	8.424	8.328	8.528	504.270	500.000	0.9
Decachlorobiphenyl	10.245	10.152	10.352	50.120	50.000	0.2
Tetrachloro-m-xylene	4.523	4.424	4.624	48.210	50.000	-3.6

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1573 SAS No.: Q1573 SDG NO.: Q1573

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL01 Date Analyzed: 03/17/2025

Lab Sample No.: AR1660CCC500 Data File : PP070589.D Time Analyzed: 14:00

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.913	4.817	5.017	513.500	500.000	2.7
Aroclor-1016-2	4.932	4.836	5.036	494.410	500.000	-1.1
Aroclor-1016-3	5.109	5.013	5.213	506.870	500.000	1.4
Aroclor-1016-4	5.150	5.054	5.254	516.010	500.000	3.2
Aroclor-1016-5	5.365	5.270	5.470	506.660	500.000	1.3
Aroclor-1260-1	6.402	6.307	6.507	504.610	500.000	0.9
Aroclor-1260-2	6.590	6.495	6.695	495.670	500.000	-0.9
Aroclor-1260-3	6.744	6.649	6.849	493.220	500.000	-1.4
Aroclor-1260-4	7.215	7.121	7.321	472.030	500.000	-5.6
Aroclor-1260-5	7.456	7.362	7.562	467.540	500.000	-6.5
Decachlorobiphenyl	8.874	8.781	8.981	46.020	50.000	-8.0
Tetrachloro-m-xylene	3.826	3.729	3.929	51.360	50.000	2.7