

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1573 SAS No.: Q1573 SDG NO.: Q1573

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 03/17/2025

Lab Sample No.: AR1660CCC500 Data File : PP070604.D Time Analyzed: 18:37

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.675	5.577	5.777	469.500	500.000	-6.1
Aroclor-1016-2	5.697	5.599	5.799	514.170	500.000	2.8
Aroclor-1016-3	5.760	5.661	5.861	487.100	500.000	-2.6
Aroclor-1016-4	5.857	5.758	5.958	482.110	500.000	-3.6
Aroclor-1016-5	6.150	6.051	6.251	481.210	500.000	-3.8
Aroclor-1260-1	7.268	7.171	7.371	506.790	500.000	1.4
Aroclor-1260-2	7.522	7.425	7.625	510.190	500.000	2.0
Aroclor-1260-3	7.880	7.783	7.983	511.570	500.000	2.3
Aroclor-1260-4	8.105	8.008	8.208	492.860	500.000	-1.4
Aroclor-1260-5	8.425	8.328	8.528	503.040	500.000	0.6
Decachlorobiphenyl	10.243	10.152	10.352	49.810	50.000	-0.4
Tetrachloro-m-xylene	4.524	4.424	4.624	49.030	50.000	-1.9

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1573 SAS No.: Q1573 SDG NO.: Q1573

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 03/17/2025

Lab Sample No.: AR1660CCC500 Data File : PP070604.D Time Analyzed: 18:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.911	4.817	5.017	510.710	500.000	2.1
Aroclor-1016-2	4.930	4.836	5.036	494.880	500.000	-1.0
Aroclor-1016-3	5.107	5.013	5.213	500.540	500.000	0.1
Aroclor-1016-4	5.149	5.054	5.254	510.540	500.000	2.1
Aroclor-1016-5	5.364	5.270	5.470	504.900	500.000	1.0
Aroclor-1260-1	6.400	6.307	6.507	489.900	500.000	-2.0
Aroclor-1260-2	6.588	6.495	6.695	481.780	500.000	-3.6
Aroclor-1260-3	6.742	6.649	6.849	478.610	500.000	-4.3
Aroclor-1260-4	7.214	7.121	7.321	459.970	500.000	-8.0
Aroclor-1260-5	7.455	7.362	7.562	464.860	500.000	-7.0
Decachlorobiphenyl	8.871	8.781	8.981	44.050	50.000	-11.9
Tetrachloro-m-xylene	3.824	3.729	3.929	49.130	50.000	-1.7