

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1573 SAS No.: Q1573 SDG NO.: Q1573

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL03 Date Analyzed: 03/17/2025

Lab Sample No.: AR1660CCC500 Data File : PP070619.D Time Analyzed: 23:13

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.671	5.577	5.777	479.620	500.000	-4.1
Aroclor-1016-2	5.693	5.599	5.799	505.960	500.000	1.2
Aroclor-1016-3	5.755	5.661	5.861	488.720	500.000	-2.3
Aroclor-1016-4	5.852	5.758	5.958	482.070	500.000	-3.6
Aroclor-1016-5	6.145	6.051	6.251	479.000	500.000	-4.2
Aroclor-1260-1	7.263	7.171	7.371	532.850	500.000	6.6
Aroclor-1260-2	7.517	7.425	7.625	519.140	500.000	3.8
Aroclor-1260-3	7.876	7.783	7.983	516.370	500.000	3.3
Aroclor-1260-4	8.100	8.008	8.208	495.460	500.000	-0.9
Aroclor-1260-5	8.421	8.328	8.528	507.150	500.000	1.4
Decachlorobiphenyl	10.238	10.152	10.352	49.870	50.000	-0.3
Tetrachloro-m-xylene	4.519	4.424	4.624	49.180	50.000	-1.6

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1573 SAS No.: Q1573 SDG NO.: Q1573

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL03 Date Analyzed: 03/17/2025

Lab Sample No.: AR1660CCC500 Data File : PP070619.D Time Analyzed: 23:13

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.909	4.817	5.017	513.180	500.000	2.6
Aroclor-1016-2	4.928	4.836	5.036	497.390	500.000	-0.5
Aroclor-1016-3	5.106	5.013	5.213	503.780	500.000	0.8
Aroclor-1016-4	5.147	5.054	5.254	499.050	500.000	-0.2
Aroclor-1016-5	5.362	5.270	5.470	514.880	500.000	3.0
Aroclor-1260-1	6.398	6.307	6.507	491.090	500.000	-1.8
Aroclor-1260-2	6.586	6.495	6.695	484.330	500.000	-3.1
Aroclor-1260-3	6.740	6.649	6.849	474.160	500.000	-5.2
Aroclor-1260-4	7.212	7.121	7.321	465.710	500.000	-6.9
Aroclor-1260-5	7.453	7.362	7.562	469.060	500.000	-6.2
Decachlorobiphenyl	8.869	8.781	8.981	46.790	50.000	-6.4
Tetrachloro-m-xylene	3.823	3.729	3.929	51.300	50.000	2.6