

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1573 SAS No.: Q1573 SDG NO.: Q1573

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL04 Date Analyzed: 03/18/2025

Lab Sample No.: AR1660CCC500 Data File : PP070632.D Time Analyzed: 03:16

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.672	5.577	5.777	477.300	500.000	-4.5
Aroclor-1016-2	5.693	5.599	5.799	504.430	500.000	0.9
Aroclor-1016-3	5.756	5.661	5.861	489.900	500.000	-2.0
Aroclor-1016-4	5.853	5.758	5.958	485.900	500.000	-2.8
Aroclor-1016-5	6.146	6.051	6.251	473.720	500.000	-5.3
Aroclor-1260-1	7.264	7.171	7.371	516.310	500.000	3.3
Aroclor-1260-2	7.517	7.425	7.625	503.450	500.000	0.7
Aroclor-1260-3	7.876	7.783	7.983	500.910	500.000	0.2
Aroclor-1260-4	8.100	8.008	8.208	475.030	500.000	-5.0
Aroclor-1260-5	8.420	8.328	8.528	494.600	500.000	-1.1
Decachlorobiphenyl	10.239	10.152	10.352	49.060	50.000	-1.9
Tetrachloro-m-xylene	4.520	4.424	4.624	48.860	50.000	-2.3

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1573 SAS No.: Q1573 SDG NO.: Q1573

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL04 Date Analyzed: 03/18/2025

Lab Sample No.: AR1660CCC500 Data File : PP070632.D Time Analyzed: 03:16

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.910	4.817	5.017	496.870	500.000	-0.6
Aroclor-1016-2	4.929	4.836	5.036	485.020	500.000	-3.0
Aroclor-1016-3	5.106	5.013	5.213	493.700	500.000	-1.3
Aroclor-1016-4	5.147	5.054	5.254	491.490	500.000	-1.7
Aroclor-1016-5	5.362	5.270	5.470	528.450	500.000	5.7
Aroclor-1260-1	6.398	6.307	6.507	484.870	500.000	-3.0
Aroclor-1260-2	6.586	6.495	6.695	472.970	500.000	-5.4
Aroclor-1260-3	6.740	6.649	6.849	467.550	500.000	-6.5
Aroclor-1260-4	7.212	7.121	7.321	445.750	500.000	-10.9
Aroclor-1260-5	7.452	7.362	7.562	445.710	500.000	-10.9
Decachlorobiphenyl	8.869	8.781	8.981	45.550	50.000	-8.9
Tetrachloro-m-xylene	3.823	3.729	3.929	50.770	50.000	1.5