

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1573 SAS No.: Q1573 SDG NO.: Q1573

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL05 Date Analyzed: 03/18/2025

Lab Sample No.: AR1660CCC500 Data File : PP070638.D Time Analyzed: 08:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.676	5.577	5.777	491.070	500.000	-1.8
Aroclor-1016-2	5.697	5.599	5.799	526.430	500.000	5.3
Aroclor-1016-3	5.759	5.661	5.861	500.210	500.000	0.0
Aroclor-1016-4	5.857	5.758	5.958	501.550	500.000	0.3
Aroclor-1016-5	6.149	6.051	6.251	507.300	500.000	1.5
Aroclor-1260-1	7.268	7.171	7.371	540.250	500.000	8.1
Aroclor-1260-2	7.521	7.425	7.625	544.090	500.000	8.8
Aroclor-1260-3	7.880	7.783	7.983	538.320	500.000	7.7
Aroclor-1260-4	8.104	8.008	8.208	509.930	500.000	2.0
Aroclor-1260-5	8.425	8.328	8.528	516.460	500.000	3.3
Decachlorobiphenyl	10.245	10.152	10.352	52.050	50.000	4.1
Tetrachloro-m-xylene	4.523	4.424	4.624	51.150	50.000	2.3

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1573 SAS No.: Q1573 SDG NO.: Q1573

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL05 Date Analyzed: 03/18/2025

Lab Sample No.: AR1660CCC500 Data File : PP070638.D Time Analyzed: 08:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.912	4.817	5.017	517.050	500.000	3.4
Aroclor-1016-2	4.930	4.836	5.036	504.670	500.000	0.9
Aroclor-1016-3	5.107	5.013	5.213	516.450	500.000	3.3
Aroclor-1016-4	5.149	5.054	5.254	505.300	500.000	1.1
Aroclor-1016-5	5.364	5.270	5.470	587.100	500.000	17.4
Aroclor-1260-1	6.399	6.307	6.507	504.340	500.000	0.9
Aroclor-1260-2	6.588	6.495	6.695	502.900	500.000	0.6
Aroclor-1260-3	6.741	6.649	6.849	478.330	500.000	-4.3
Aroclor-1260-4	7.213	7.121	7.321	474.490	500.000	-5.1
Aroclor-1260-5	7.454	7.362	7.562	490.680	500.000	-1.9
Decachlorobiphenyl	8.870	8.781	8.981	49.050	50.000	-1.9
Tetrachloro-m-xylene	3.825	3.729	3.929	49.410	50.000	-1.2