

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1573 SAS No.: Q1573 SDG NO.: Q1573

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL06 Date Analyzed: 03/18/2025

Lab Sample No.: AR1660CCC500 Data File : PP070647.D Time Analyzed: 11:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.672	5.577	5.777	475.020	500.000	-5.0
Aroclor-1016-2	5.694	5.599	5.799	517.610	500.000	3.5
Aroclor-1016-3	5.756	5.661	5.861	490.300	500.000	-1.9
Aroclor-1016-4	5.853	5.758	5.958	476.340	500.000	-4.7
Aroclor-1016-5	6.146	6.051	6.251	499.370	500.000	-0.1
Aroclor-1260-1	7.264	7.171	7.371	569.940	500.000	14.0
Aroclor-1260-2	7.518	7.425	7.625	541.090	500.000	8.2
Aroclor-1260-3	7.876	7.783	7.983	533.210	500.000	6.6
Aroclor-1260-4	8.101	8.008	8.208	506.300	500.000	1.3
Aroclor-1260-5	8.421	8.328	8.528	514.580	500.000	2.9
Decachlorobiphenyl	10.238	10.152	10.352	50.900	50.000	1.8
Tetrachloro-m-xylene	4.521	4.424	4.624	49.830	50.000	-0.3

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: Q1573 SAS No.: Q1573 SDG NO.: Q1573

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL06 Date Analyzed: 03/18/2025

Lab Sample No.: AR1660CCC500 Data File : PP070647.D Time Analyzed: 11:19

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.909	4.817	5.017	493.750	500.000	-1.3
Aroclor-1016-2	4.928	4.836	5.036	487.120	500.000	-2.6
Aroclor-1016-3	5.105	5.013	5.213	490.310	500.000	-1.9
Aroclor-1016-4	5.147	5.054	5.254	481.210	500.000	-3.8
Aroclor-1016-5	5.361	5.270	5.470	508.140	500.000	1.6
Aroclor-1260-1	6.398	6.307	6.507	511.740	500.000	2.3
Aroclor-1260-2	6.585	6.495	6.695	490.830	500.000	-1.8
Aroclor-1260-3	6.739	6.649	6.849	476.910	500.000	-4.6
Aroclor-1260-4	7.211	7.121	7.321	466.460	500.000	-6.7
Aroclor-1260-5	7.452	7.362	7.562	462.540	500.000	-7.5
Decachlorobiphenyl	8.868	8.781	8.981	47.240	50.000	-5.5
Tetrachloro-m-xylene	3.823	3.729	3.929	50.280	50.000	0.6