

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP031125

Review By	yogesh	Review On	3/12/2025 7:57:33 AM
Supervise By	mohammad	Supervise On	3/19/2025 6:54:29 AM
SubDirectory	PP031125	HP Acquire Method	HP Processing Method PP031125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773
Internal Standard/PEM	
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP070417.D	11 Mar 2025 14:41		YPIAJ	Ok
2	I.BLK	I.BLK	PP070418.D	11 Mar 2025 14:57		YPIAJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PP070419.D	11 Mar 2025 15:13		YPIAJ	Ok
4	AR1660ICC750	AR1660ICC750	PP070420.D	11 Mar 2025 15:29		YPIAJ	Ok
5	AR1660ICC500	AR1660ICC500	PP070421.D	11 Mar 2025 15:46		YPIAJ	Ok
6	AR1660ICC250	AR1660ICC250	PP070422.D	11 Mar 2025 16:02		YPIAJ	Ok,M
7	AR1660ICC050	AR1660ICC050	PP070423.D	11 Mar 2025 16:18		YPIAJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP070424.D	11 Mar 2025 16:34		YPIAJ	Ok
9	AR1232ICC1000	AR1232ICC1000	PP070425.D	11 Mar 2025 16:51		YPIAJ	Ok
10	AR1232ICC750	AR1232ICC750	PP070426.D	11 Mar 2025 17:07		YPIAJ	Ok
11	AR1232ICC500	AR1232ICC500	PP070427.D	11 Mar 2025 17:23		YPIAJ	Ok
12	AR1232ICC250	AR1232ICC250	PP070428.D	11 Mar 2025 17:40		YPIAJ	Ok,M
13	AR1232ICC050	AR1232ICC050	PP070429.D	11 Mar 2025 17:56		YPIAJ	Ok,M
14	AR1242ICC1000	AR1242ICC1000	PP070430.D	11 Mar 2025 18:12		YPIAJ	Ok
15	AR1242ICC750	AR1242ICC750	PP070431.D	11 Mar 2025 18:28		YPIAJ	Ok
16	AR1242ICC500	AR1242ICC500	PP070432.D	11 Mar 2025 18:45		YPIAJ	Ok
17	AR1242ICC250	AR1242ICC250	PP070433.D	11 Mar 2025 19:01		YPIAJ	Ok
18	AR1242ICC050	AR1242ICC050	PP070434.D	11 Mar 2025 19:17		YPIAJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP031125

Review By	yogesh	Review On	3/12/2025 7:57:33 AM
Supervise By	mohammad	Supervise On	3/19/2025 6:54:29 AM
SubDirectory	PP031125	HP Acquire Method	HP Processing Method PP031125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248ICC1000	AR1248ICC1000	PP070435.D	11 Mar 2025 19:34		YPIAJ	Ok
20	AR1248ICC750	AR1248ICC750	PP070436.D	11 Mar 2025 19:50		YPIAJ	Ok
21	AR1248ICC500	AR1248ICC500	PP070437.D	11 Mar 2025 20:06		YPIAJ	Ok
22	AR1248ICC250	AR1248ICC250	PP070438.D	11 Mar 2025 20:22		YPIAJ	Ok
23	AR1248ICC050	AR1248ICC050	PP070439.D	11 Mar 2025 20:39		YPIAJ	Ok,M
24	AR1254ICC1000	AR1254ICC1000	PP070440.D	11 Mar 2025 20:55		YPIAJ	Ok
25	AR1254ICC750	AR1254ICC750	PP070441.D	11 Mar 2025 21:11		YPIAJ	Ok
26	AR1254ICC500	AR1254ICC500	PP070442.D	11 Mar 2025 21:28		YPIAJ	Ok
27	AR1254ICC250	AR1254ICC250	PP070443.D	11 Mar 2025 21:44		YPIAJ	Ok
28	AR1254ICC050	AR1254ICC050	PP070444.D	11 Mar 2025 22:00		YPIAJ	Ok
29	AR1262ICC500	AR1262ICC500	PP070445.D	11 Mar 2025 22:16		YPIAJ	Ok
30	AR1268ICC1000	AR1268ICC1000	PP070446.D	11 Mar 2025 22:33		YPIAJ	Ok
31	AR1268ICC750	AR1268ICC750	PP070447.D	11 Mar 2025 22:49		YPIAJ	Ok
32	AR1268ICC500	AR1268ICC500	PP070448.D	11 Mar 2025 23:05		YPIAJ	Ok
33	AR1268ICC250	AR1268ICC250	PP070449.D	11 Mar 2025 23:22		YPIAJ	Ok
34	AR1268ICC050	AR1268ICC050	PP070450.D	11 Mar 2025 23:38		YPIAJ	Ok,M
35	PP031125ICV500	ICVPP031125	PP070451.D	11 Mar 2025 23:54		YPIAJ	Ok
36	AR1232ICV500	ICVPP031125AR1232	PP070452.D	12 Mar 2025 00:10		YPIAJ	Ok
37	AR1242ICV500	ICVPP031125AR1242	PP070453.D	12 Mar 2025 00:27		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP031125

Review By	yogesh	Review On	3/12/2025 7:57:33 AM
Supervise By	mohammad	Supervise On	3/19/2025 6:54:29 AM
SubDirectory	PP031125	HP Acquire Method	HP Processing Method PP031125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	AR1248ICV500	ICVPP031125AR1248	PP070454.D	12 Mar 2025 00:43		YPIAJ	Ok
39	AR1254ICV500	ICVPP031125AR1254	PP070455.D	12 Mar 2025 00:59		YPIAJ	Ok
40	AR1268ICV500	ICVPP031125AR1268	PP070456.D	12 Mar 2025 01:16		YPIAJ	Ok
41	AR1660CCC500	AR1660CCC500	PP070457.D	12 Mar 2025 01:32		YPIAJ	Ok
42	AR1232CCC500	AR1232CCC500	PP070458.D	12 Mar 2025 01:48		YPIAJ	Ok
43	AR1242CCC500	AR1242CCC500	PP070459.D	12 Mar 2025 02:04		YPIAJ	Ok
44	I.BLK	I.BLK	PP070460.D	12 Mar 2025 02:21		YPIAJ	Ok
45	PB167066BL	PB167066BL	PP070461.D	12 Mar 2025 02:37	Already run	YPIAJ	Not Ok
46	Q1502-15	PT-PCBW-WP	PP070462.D	12 Mar 2025 02:53	AR1232 HIT	YPIAJ	Ok
47	AR1660CCC500	AR1660CCC500	PP070463.D	12 Mar 2025 03:09		YPIAJ	Ok
48	AR1232CCC500	AR1232CCC500	PP070464.D	12 Mar 2025 03:26		YPIAJ	Ok
49	AR1242CCC500	AR1242CCC500	PP070465.D	12 Mar 2025 03:42		YPIAJ	Ok,M
50	I.BLK	I.BLK	PP070466.D	12 Mar 2025 03:58		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP031725

Review By	yogesh	Review On	3/17/2025 11:38:20 AM
Supervise By	mohammad	Supervise On	3/19/2025 6:47:02 AM
SubDirectory	PP031725	HP Acquire Method	HP Processing Method PP031125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP070573.D	17 Mar 2025 08:51		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PP070574.D	17 Mar 2025 09:08		YP\AJ	Ok
3	AR1242CCC500	AR1242CCC500	PP070575.D	17 Mar 2025 09:25		YP\AJ	Ok
4	AR1248CCC500	AR1248CCC500	PP070576.D	17 Mar 2025 09:41		YP\AJ	Ok
5	AR1254CCC500	AR1254CCC500	PP070577.D	17 Mar 2025 09:57		YP\AJ	Ok
6	I.BLK	I.BLK	PP070578.D	17 Mar 2025 10:13		YP\AJ	Ok
7	Q1572-01DL	C0AC9DL	PP070579.D	17 Mar 2025 10:30	AR1248 + AR1260 Hit, Need Dilution, DCB high in 2nd column	YP\AJ	Dilution
8	Q1572-01DL2	C0AC9DL2	PP070580.D	17 Mar 2025 10:46	AR1248 + AR1260 Hit	YP\AJ	Ok
9	Q1572-02DL	C0O14DL	PP070581.D	17 Mar 2025 11:02	AR1260 + AR1248 Hit , Need dilution, DCB high 2nd column, TCMX high in Both column	YP\AJ	Dilution
10	Q1572-02DL2	C0O14DL2	PP070582.D	17 Mar 2025 11:18	AR1260 AR1248 Hit	YP\AJ	Ok
11	Q1572-03DL	C0O15DL	PP070583.D	17 Mar 2025 11:35	AR1260 AR1248 Hit , Need dilution	YP\AJ	Dilution
12	Q1572-03DL2	C0O15DL2	PP070584.D	17 Mar 2025 11:51	AR1260 AR1248 Hit	YP\AJ	Ok
13	Q1572-04DL	C0O16DL	PP070585.D	17 Mar 2025 12:07	AR1260 + AR1248 Hit, Need Dilution	YP\AJ	Dilution
14	Q1572-04DL2	C0O16DL2	PP070586.D	17 Mar 2025 12:23	AR1260 + AR1248 Hit	YP\AJ	Ok,NR
15	Q1572-05DL	C0O17DL	PP070587.D	17 Mar 2025 12:40	AR1248 + AR1260 Hit, Need Dilution, DCB High in 2nd column	YP\AJ	Dilution
16	Q1572-05DL2	C0O17DL2	PP070588.D	17 Mar 2025 12:56	AR1248 + AR1260 Hit,	YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP031725

Review By	yogesh	Review On	3/17/2025 11:38:20 AM
Supervise By	mohammad	Supervise On	3/19/2025 6:47:02 AM
SubDirectory	PP031725	HP Acquire Method	HP Processing Method PP031125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

17	AR1660CCC500	AR1660CCC500	PP070589.D	17 Mar 2025 14:00		YPIAJ	Ok
18	AR1242CCC500	AR1242CCC500	PP070590.D	17 Mar 2025 14:17		YPIAJ	Ok
19	AR1248CCC500	AR1248CCC500	PP070591.D	17 Mar 2025 14:33		YPIAJ	Ok
20	AR1254CCC500	AR1254CCC500	PP070592.D	17 Mar 2025 14:49		YPIAJ	Ok
21	I.BLK	I.BLK	PP070593.D	17 Mar 2025 15:05		YPIAJ	Ok
22	PB167154BL	PB167154BL	PP070594.D	17 Mar 2025 15:22		YPIAJ	Ok
23	PB167154BS	PB167154BS	PP070595.D	17 Mar 2025 15:38		YPIAJ	Ok
24	Q1572-07DL	C0O19DL	PP070596.D	17 Mar 2025 15:54	AR1260 AR1248 Hit	YPIAJ	Ok,M
25	Q1572-06DL	C0O18DL	PP070597.D	17 Mar 2025 16:10	AR1260 AR1248 Hit, Need Further 20X Dilution	YPIAJ	Dilution
26	Q1572-06DL2	C0O18DL2	PP070598.D	17 Mar 2025 16:27	AR1260 AR1248 Hit, Need Further 2X Dilution	YPIAJ	Not Ok
27	Q1572-08DL	C0O20DL	PP070599.D	17 Mar 2025 16:43	AR1248 Hit, DCB High in 2nd column	YPIAJ	Ok,M
28	Q1572-09DL	C0O21DL	PP070600.D	17 Mar 2025 16:59	AR1248 Hit	YPIAJ	Ok
29	Q1573-01	C0AC4	PP070601.D	17 Mar 2025 17:15	AR1248 + AR1260 Hit	YPIAJ	Ok,M
30	Q1573-01MS	C0AC4MS	PP070602.D	17 Mar 2025 17:32		YPIAJ	Ok,M
31	Q1573-01MSD	C0AC4MSD	PP070603.D	17 Mar 2025 17:48		YPIAJ	Ok,M
32	AR1660CCC500	AR1660CCC500	PP070604.D	17 Mar 2025 18:37		YPIAJ	Ok
33	AR1242CCC500	AR1242CCC500	PP070605.D	17 Mar 2025 18:53		YPIAJ	Ok
34	AR1248CCC500	AR1248CCC500	PP070606.D	17 Mar 2025 19:09		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP031725

Review By	yogesh	Review On	3/17/2025 11:38:20 AM
Supervise By	mohammad	Supervise On	3/19/2025 6:47:02 AM
SubDirectory	PP031725	HP Acquire Method	HP Processing Method PP031125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

35	AR1254CCC500	AR1254CCC500	PP070607.D	17 Mar 2025 19:25		YPIAJ	Ok
36	I.BLK	I.BLK	PP070608.D	17 Mar 2025 19:42		YPIAJ	Ok
37	Q1573-02	C0AC5	PP070609.D	17 Mar 2025 19:58	AR1248 + AR1260 Hit, DCB Low in 1st column	YPIAJ	Ok,M
38	Q1573-03	C0AC6	PP070610.D	17 Mar 2025 20:14	AR1248 + AR1260 Hit, DCB Low in both column	YPIAJ	ReRun
39	Q1573-04	C0AC8	PP070611.D	17 Mar 2025 20:30	AR1248 + AR1260 Hit	YPIAJ	Ok,M
40	Q1573-05	C0O29	PP070612.D	17 Mar 2025 20:47	AR1260 Hit	YPIAJ	Ok,M
41	Q1573-06	C0O30	PP070613.D	17 Mar 2025 21:03	AR1260 Hit	YPIAJ	Ok
42	Q1573-07	C0O31	PP070614.D	17 Mar 2025 21:19	AR1248 + AR1260 Hit	YPIAJ	Ok,M
43	Q1573-08	C0O32	PP070615.D	17 Mar 2025 21:35	AR1248 + AR1260 Hit	YPIAJ	Ok,M
44	Q1573-09	C0O33	PP070616.D	17 Mar 2025 21:52	AR1248 + AR1260 Hit	YPIAJ	Ok,M
45	Q1573-10	C0O34	PP070617.D	17 Mar 2025 22:08	AR1248 + AR1260 Hit	YPIAJ	Ok,M
46	Q1573-11	C0O35	PP070618.D	17 Mar 2025 22:24	AR1248 + AR1260 Hit, DCB low in both column	YPIAJ	ReRun
47	AR1660CCC500	AR1660CCC500	PP070619.D	17 Mar 2025 23:13		YPIAJ	Ok
48	AR1242CCC500	AR1242CCC500	PP070620.D	17 Mar 2025 23:29		YPIAJ	Ok
49	AR1248CCC500	AR1248CCC500	PP070621.D	17 Mar 2025 23:45		YPIAJ	Ok
50	AR1254CCC500	AR1254CCC500	PP070622.D	18 Mar 2025 00:02		YPIAJ	Ok
51	I.BLK	I.BLK	PP070623.D	18 Mar 2025 00:18		YPIAJ	Ok
52	Q1581-01	TR-06-031425	PP070624.D	18 Mar 2025 00:34		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP031725

Review By	yogesh	Review On	3/17/2025 11:38:20 AM
Supervise By	mohammad	Supervise On	3/19/2025 6:47:02 AM
SubDirectory	PP031725	HP Acquire Method	HP Processing Method PP031125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

53	Q1585-01	OK-02-03142025	PP070625.D	18 Mar 2025 00:50		YPIAJ	Ok,M
54	PB167172BL	PB167172BL	PP070626.D	18 Mar 2025 01:06		YPIAJ	Ok
55	PB167172BS	PB167172BS	PP070627.D	18 Mar 2025 01:23		YPIAJ	Ok
56	PB167172BSD	PB167172BSD	PP070628.D	18 Mar 2025 01:39		YPIAJ	Ok,M
57	Q1573-12	C0AC7	PP070629.D	18 Mar 2025 01:55		YPIAJ	Ok
58	Q1573-13	C0O37	PP070630.D	18 Mar 2025 02:11		YPIAJ	Ok
59	Q1577-03	MOO-25-0072	PP070631.D	18 Mar 2025 02:28		YPIAJ	Ok,M
60	AR1660CCC500	AR1660CCC500	PP070632.D	18 Mar 2025 03:16		YPIAJ	Ok
61	AR1242CCC500	AR1242CCC500	PP070633.D	18 Mar 2025 03:32		YPIAJ	Ok
62	AR1248CCC500	AR1248CCC500	PP070634.D	18 Mar 2025 03:49		YPIAJ	Ok
63	AR1254CCC500	AR1254CCC500	PP070635.D	18 Mar 2025 04:05		YPIAJ	Ok
64	I.BLK	I.BLK	PP070636.D	18 Mar 2025 04:21		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP031825

Review By	yogesh	Review On	3/18/2025 9:12:10 AM
Supervise By	mohammad	Supervise On	3/24/2025 3:02:25 AM
SubDirectory	PP031825	HP Acquire Method	HP Processing Method PP031125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773
Internal Standard/PEM	
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP070637.D	18 Mar 2025 07:48		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PP070638.D	18 Mar 2025 08:04		YP\AJ	Ok
3	AR1242CCC500	AR1242CCC500	PP070639.D	18 Mar 2025 08:21		YP\AJ	Ok
4	AR1248CCC500	AR1248CCC500	PP070640.D	18 Mar 2025 08:37		YP\AJ	Ok
5	AR1254CCC500	AR1254CCC500	PP070641.D	18 Mar 2025 08:53		YP\AJ	Ok,M
6	I.BLK	I.BLK	PP070642.D	18 Mar 2025 09:10		YP\AJ	Ok,M
7	PP24328	PP24328	PP070643.D	18 Mar 2025 09:26		YP\AJ	Ok
8	Q1573-03	C0AC6	PP070644.D	18 Mar 2025 09:42	AR1248 + AR1260 Hit	YP\AJ	Ok,M
9	Q1573-11	C0O35	PP070645.D	18 Mar 2025 10:14	AR1248 + AR1260 Hit	YP\AJ	Ok,M
10	Q1572-06DL2	C0O18DL2	PP070646.D	18 Mar 2025 10:31	AR1260 AR1248 Hit,	YP\AJ	Ok,M
11	AR1660CCC500	AR1660CCC500	PP070647.D	18 Mar 2025 11:19		YP\AJ	Ok,M
12	AR1242CCC500	AR1242CCC500	PP070648.D	18 Mar 2025 11:36		YP\AJ	Ok,M
13	AR1248CCC500	AR1248CCC500	PP070649.D	18 Mar 2025 11:52		YP\AJ	Ok
14	AR1254CCC500	AR1254CCC500	PP070650.D	18 Mar 2025 12:08		YP\AJ	Ok,M
15	I.BLK	I.BLK	PP070651.D	18 Mar 2025 12:24		YP\AJ	Ok
16	PB167186BL	PB167186BL	PP070652.D	18 Mar 2025 15:02		YP\AJ	Ok
17	PB167186BS	PB167186BS	PP070653.D	18 Mar 2025 15:18		YP\AJ	Ok
18	Q1591-02	GAS-TRE-25-0037	PP070654.D	18 Mar 2025 15:35	AR1254 Hit	YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP031825

Review By	yogesh	Review On	3/18/2025 9:12:10 AM
Supervise By	mohammad	Supervise On	3/24/2025 3:02:25 AM
SubDirectory	PP031825	HP Acquire Method	HP Processing Method PP031125
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775 PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773 PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		

19	Q1591-03	ARS20-0008	PP070655.D	18 Mar 2025 15:51		YPIAJ	Ok
20	Q1591-03MS	ARS20-0008MS	PP070656.D	18 Mar 2025 16:07		YPIAJ	Ok,M
21	Q1591-03MSD	ARS20-0008MSD	PP070657.D	18 Mar 2025 16:23		YPIAJ	Ok,M
22	Q1592-01	OILY-DEBRIS-COMP	PP070658.D	18 Mar 2025 16:40		YPIAJ	Ok
23	Q1590-01	3794	PP070659.D	18 Mar 2025 16:56		YPIAJ	Ok
24	Q1590-02	20071	PP070660.D	18 Mar 2025 17:12		YPIAJ	Ok,M
25	Q1591-01	GAS-TRE-1084	PP070661.D	18 Mar 2025 17:28		YPIAJ	Ok,M
26	AR1660CCC500	AR1660CCC500	PP070662.D	18 Mar 2025 18:17		YPIAJ	Ok,M
27	AR1242CCC500	AR1242CCC500	PP070663.D	18 Mar 2025 18:34		YPIAJ	Ok
28	AR1248CCC500	AR1248CCC500	PP070664.D	18 Mar 2025 18:50		YPIAJ	Ok
29	AR1254CCC500	AR1254CCC500	PP070665.D	18 Mar 2025 19:06		YPIAJ	Ok,M
30	I.BLK	I.BLK	PP070666.D	18 Mar 2025 19:22		YPIAJ	Ok
31	PB167184BL	PB167184BL	PP070667.D	18 Mar 2025 19:39		YPIAJ	Ok
32	PB167184BS	PB167184BS	PP070668.D	18 Mar 2025 19:55		YPIAJ	Ok
33	PB167184BSD	PB167184BSD	PP070669.D	18 Mar 2025 20:11		YPIAJ	Ok
34	Q1591-04	GAS-TRE-25-0026	PP070670.D	18 Mar 2025 20:27		YPIAJ	Ok
35	Q1591-05	GAS-TRE-25-0027	PP070671.D	18 Mar 2025 20:44		YPIAJ	Ok,M
36	Q1591-06	GAS-TRE-25-0028	PP070672.D	18 Mar 2025 21:00		YPIAJ	Ok,M
37	Q1592-03	257952	PP070673.D	18 Mar 2025 21:16		YPIAJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP031825

Review By	yogesh	Review On	3/18/2025 9:12:10 AM
Supervise By	mohammad	Supervise On	3/24/2025 3:02:25 AM
SubDirectory	PP031825	HP Acquire Method	HP Processing Method PP031125
STD. NAME	STD REF.#		
Tune/Reschk	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775 PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773 PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Initial Calibration Stds			
CCC			
Internal Standard/PEM			
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	Q1593-01	258829	PP070674.D	18 Mar 2025 21:32		YPIAJ	Ok
39	Q1594-01	GAS-CAM-40615	PP070675.D	18 Mar 2025 21:49		YPIAJ	Ok
40	Q1590-04	3799	PP070676.D	18 Mar 2025 22:05		YPIAJ	Ok,M
41	AR1660CCC500	AR1660CCC500	PP070677.D	18 Mar 2025 22:54		YPIAJ	Ok
42	AR1242CCC500	AR1242CCC500	PP070678.D	18 Mar 2025 23:10		YPIAJ	Ok,M
43	AR1248CCC500	AR1248CCC500	PP070679.D	18 Mar 2025 23:26		YPIAJ	Ok
44	AR1254CCC500	AR1254CCC500	PP070680.D	18 Mar 2025 23:42		YPIAJ	Ok,M
45	I.BLK	I.BLK	PP070681.D	18 Mar 2025 23:59		YPIAJ	Ok

M : Manual Integration