

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1573

Client: Tetra Tech, EMI

Analytical Method: 8082A

DataFile : PP070603.D

Lab Sample ID:	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
			Result	Result			Qual	RPD	Qual	Low	High	RPD
Client Sample ID: Q1573-01MSD	C0AC4MSD											
	AR1016	229.7	0	239	ug/kg	104		3		55	146	20
	AR1260	229.7	23.7	198	ug/kg	76		1		31	146	20