



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1573

Client: Tetra Tech, EMI

Analytical Method: **8082A** **Datafile :** PP070628.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB167172BSD	AR1016	5	4.50	ug/L	90	0			61	112	20
	AR1260	5	4.40	ug/L	88	5			66	113	20