



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: G Environmental **SDG No.:** Q1574
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q1574 **SAS No.:** Q1574
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.19	0.2	97	40 - 160	CV	03/17/2025	13:37	LB135056
CRI01	Aluminum	106	100	106	40 - 160	P	03/18/2025	10:49	LB135083
	Antimony	53.8	50.0	108	40 - 160	P	03/18/2025	10:49	LB135083
	Arsenic	22.6	20.0	113	40 - 160	P	03/18/2025	10:49	LB135083
	Barium	94.4	100	94	40 - 160	P	03/18/2025	10:49	LB135083
	Beryllium	5.87	6.0	98	40 - 160	P	03/18/2025	10:49	LB135083
	Cadmium	5.91	6.0	98	40 - 160	P	03/18/2025	10:49	LB135083
	Calcium	1960	2000	98	40 - 160	P	03/18/2025	10:49	LB135083
	Chromium	10.8	10.0	108	40 - 160	P	03/18/2025	10:49	LB135083
	Cobalt	29.4	30.0	98	40 - 160	P	03/18/2025	10:49	LB135083
	Copper	22.0	20.0	110	40 - 160	P	03/18/2025	10:49	LB135083
	Iron	114	100	114	40 - 160	P	03/18/2025	10:49	LB135083
	Lead	11.3	12.0	94	40 - 160	P	03/18/2025	10:49	LB135083
	Magnesium	2060	2000	103	40 - 160	P	03/18/2025	10:49	LB135083
	Manganese	18.5	20.0	93	40 - 160	P	03/18/2025	10:49	LB135083
	Nickel	39.3	40.0	98	40 - 160	P	03/18/2025	10:49	LB135083
	Potassium	1840	2000	92	40 - 160	P	03/18/2025	10:49	LB135083
	Selenium	16.6	20.0	83	40 - 160	P	03/18/2025	10:49	LB135083
	Silver	10.7	10.0	107	40 - 160	P	03/18/2025	10:49	LB135083
	Sodium	1950	2000	97	40 - 160	P	03/18/2025	10:49	LB135083
	Thallium	38.4	40.0	96	40 - 160	P	03/18/2025	10:49	LB135083
	Vanadium	37.0	40.0	92	40 - 160	P	03/18/2025	10:49	LB135083
	Zinc	43.1	40.0	108	40 - 160	P	03/18/2025	10:49	LB135083