



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	G Environmental				SDG No.:	Q1574				
Contract:	GENV01	Lab Code:	CHEM		Case No.:	Q1574	SAS No.:	Q1574		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB45	Mercury	0.20	+/-0.20	U	0.20	CV	03/17/2025	13:30	LB135056	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q1574

Contract: GENV01

Lab Code: CHEM

Case No.: Q1574

SAS No.: Q1574

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB17	Mercury	0.20	+/-0.20	U	0.20	CV	03/17/2025	13:35	LB135056
CCB18	Mercury	0.20	+/-0.20	U	0.20	CV	03/17/2025	14:07	LB135056
CCB19	Mercury	0.20	+/-0.20	U	0.20	CV	03/17/2025	14:42	LB135056

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>G Environmental</u>		SDG No.: <u>Q1574</u>							
Contract: <u>GENV01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1574</u>	SAS No.: <u>Q1574</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	03/18/2025	10:45	LB135083
	Antimony	50.0	+/-50.0	U	50.0	P	03/18/2025	10:45	LB135083
	Arsenic	20.0	+/-20.0	U	20.0	P	03/18/2025	10:45	LB135083
	Barium	100	+/-100	U	100	P	03/18/2025	10:45	LB135083
	Beryllium	6.00	+/-6.00	U	6.00	P	03/18/2025	10:45	LB135083
	Cadmium	6.00	+/-6.00	U	6.00	P	03/18/2025	10:45	LB135083
	Calcium	2000	+/-2000	U	2000	P	03/18/2025	10:45	LB135083
	Chromium	10.0	+/-10.0	U	10.0	P	03/18/2025	10:45	LB135083
	Cobalt	30.0	+/-30.0	U	30.0	P	03/18/2025	10:45	LB135083
	Copper	20.0	+/-20.0	U	20.0	P	03/18/2025	10:45	LB135083
	Iron	100	+/-100	U	100	P	03/18/2025	10:45	LB135083
	Lead	12.0	+/-12.0	U	12.0	P	03/18/2025	10:45	LB135083
	Magnesium	2000	+/-2000	U	2000	P	03/18/2025	10:45	LB135083
	Manganese	20.0	+/-20.0	U	20.0	P	03/18/2025	10:45	LB135083
	Nickel	40.0	+/-40.0	U	40.0	P	03/18/2025	10:45	LB135083
	Potassium	2000	+/-2000	U	2000	P	03/18/2025	10:45	LB135083
	Selenium	20.0	+/-20.0	U	20.0	P	03/18/2025	10:45	LB135083
	Silver	10.0	+/-10.0	U	10.0	P	03/18/2025	10:45	LB135083
	Sodium	2000	+/-2000	U	2000	P	03/18/2025	10:45	LB135083
	Thallium	40.0	+/-40.0	U	40.0	P	03/18/2025	10:45	LB135083
	Vanadium	40.0	+/-40.0	U	40.0	P	03/18/2025	10:45	LB135083
	Zinc	40.0	+/-40.0	U	40.0	P	03/18/2025	10:45	LB135083

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental		SDG No.: Q1574							
Contract: GENV01	Lab Code: CHEM	Case No.: Q1574	SAS No.: Q1574						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	03/18/2025	11:13	LB135083
	Antimony	50.0	+/-50.0	U	50.0	P	03/18/2025	11:13	LB135083
	Arsenic	20.0	+/-20.0	U	20.0	P	03/18/2025	11:13	LB135083
	Barium	100	+/-100	U	100	P	03/18/2025	11:13	LB135083
	Beryllium	6.00	+/-6.00	U	6.00	P	03/18/2025	11:13	LB135083
	Cadmium	6.00	+/-6.00	U	6.00	P	03/18/2025	11:13	LB135083
	Calcium	2000	+/-2000	U	2000	P	03/18/2025	11:13	LB135083
	Chromium	10.0	+/-10.0	U	10.0	P	03/18/2025	11:13	LB135083
	Cobalt	30.0	+/-30.0	U	30.0	P	03/18/2025	11:13	LB135083
	Copper	20.0	+/-20.0	U	20.0	P	03/18/2025	11:13	LB135083
	Iron	100	+/-100	U	100	P	03/18/2025	11:13	LB135083
	Lead	12.0	+/-12.0	U	12.0	P	03/18/2025	11:13	LB135083
	Magnesium	2000	+/-2000	U	2000	P	03/18/2025	11:13	LB135083
	Manganese	20.0	+/-20.0	U	20.0	P	03/18/2025	11:13	LB135083
	Nickel	40.0	+/-40.0	U	40.0	P	03/18/2025	11:13	LB135083
	Potassium	2000	+/-2000	U	2000	P	03/18/2025	11:13	LB135083
	Selenium	20.0	+/-20.0	U	20.0	P	03/18/2025	11:13	LB135083
	Silver	10.0	+/-10.0	U	10.0	P	03/18/2025	11:13	LB135083
	Sodium	2000	+/-2000	U	2000	P	03/18/2025	11:13	LB135083
	Thallium	40.0	+/-40.0	U	40.0	P	03/18/2025	11:13	LB135083
	Vanadium	40.0	+/-40.0	U	40.0	P	03/18/2025	11:13	LB135083
	Zinc	40.0	+/-40.0	U	40.0	P	03/18/2025	11:13	LB135083
CCB02	Aluminum	100	+/-100	U	100	P	03/18/2025	12:17	LB135083
	Antimony	50.0	+/-50.0	U	50.0	P	03/18/2025	12:17	LB135083
	Arsenic	20.0	+/-20.0	U	20.0	P	03/18/2025	12:17	LB135083
	Barium	100	+/-100	U	100	P	03/18/2025	12:17	LB135083
	Beryllium	6.00	+/-6.00	U	6.00	P	03/18/2025	12:17	LB135083
	Cadmium	6.00	+/-6.00	U	6.00	P	03/18/2025	12:17	LB135083
	Calcium	2000	+/-2000	U	2000	P	03/18/2025	12:17	LB135083
	Chromium	10.0	+/-10.0	U	10.0	P	03/18/2025	12:17	LB135083
	Cobalt	30.0	+/-30.0	U	30.0	P	03/18/2025	12:17	LB135083
	Copper	20.0	+/-20.0	U	20.0	P	03/18/2025	12:17	LB135083
	Iron	100	+/-100	U	100	P	03/18/2025	12:17	LB135083
	Lead	12.0	+/-12.0	U	12.0	P	03/18/2025	12:17	LB135083
	Magnesium	2000	+/-2000	U	2000	P	03/18/2025	12:17	LB135083
	Manganese	20.0	+/-20.0	U	20.0	P	03/18/2025	12:17	LB135083
	Nickel	40.0	+/-40.0	U	40.0	P	03/18/2025	12:17	LB135083
	Potassium	2000	+/-2000	U	2000	P	03/18/2025	12:17	LB135083
	Selenium	20.0	+/-20.0	U	20.0	P	03/18/2025	12:17	LB135083

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>G Environmental</u>		SDG No.: <u>Q1574</u>							
Contract: <u>GENV01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1574</u>	SAS No.: <u>Q1574</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	03/18/2025	12:17	LB135083
	Sodium	2000	+/-2000	U	2000	P	03/18/2025	12:17	LB135083
	Thallium	40.0	+/-40.0	U	40.0	P	03/18/2025	12:17	LB135083
	Vanadium	40.0	+/-40.0	U	40.0	P	03/18/2025	12:17	LB135083
	Zinc	40.0	+/-40.0	U	40.0	P	03/18/2025	12:17	LB135083
CCB03	Aluminum	100	+/-100	U	100	P	03/18/2025	13:13	LB135083
	Antimony	50.0	+/-50.0	U	50.0	P	03/18/2025	13:13	LB135083
	Arsenic	20.0	+/-20.0	U	20.0	P	03/18/2025	13:13	LB135083
	Barium	100	+/-100	U	100	P	03/18/2025	13:13	LB135083
	Beryllium	6.00	+/-6.00	U	6.00	P	03/18/2025	13:13	LB135083
	Cadmium	6.00	+/-6.00	U	6.00	P	03/18/2025	13:13	LB135083
	Calcium	2000	+/-2000	U	2000	P	03/18/2025	13:13	LB135083
	Chromium	10.0	+/-10.0	U	10.0	P	03/18/2025	13:13	LB135083
	Cobalt	30.0	+/-30.0	U	30.0	P	03/18/2025	13:13	LB135083
	Copper	20.0	+/-20.0	U	20.0	P	03/18/2025	13:13	LB135083
	Iron	100	+/-100	U	100	P	03/18/2025	13:13	LB135083
	Lead	12.0	+/-12.0	U	12.0	P	03/18/2025	13:13	LB135083
	Magnesium	2000	+/-2000	U	2000	P	03/18/2025	13:13	LB135083
	Manganese	20.0	+/-20.0	U	20.0	P	03/18/2025	13:13	LB135083
	Nickel	40.0	+/-40.0	U	40.0	P	03/18/2025	13:13	LB135083
	Potassium	2000	+/-2000	U	2000	P	03/18/2025	13:13	LB135083
	Selenium	20.0	+/-20.0	U	20.0	P	03/18/2025	13:13	LB135083
	Silver	10.0	+/-10.0	U	10.0	P	03/18/2025	13:13	LB135083
	Sodium	2000	+/-2000	U	2000	P	03/18/2025	13:13	LB135083
	Thallium	40.0	+/-40.0	U	40.0	P	03/18/2025	13:13	LB135083
	Vanadium	40.0	+/-40.0	U	40.0	P	03/18/2025	13:13	LB135083
	Zinc	40.0	+/-40.0	U	40.0	P	03/18/2025	13:13	LB135083
CCB04	Aluminum	100	+/-100	U	100	P	03/18/2025	14:20	LB135083
	Antimony	50.0	+/-50.0	U	50.0	P	03/18/2025	14:20	LB135083
	Arsenic	20.0	+/-20.0	U	20.0	P	03/18/2025	14:20	LB135083
	Barium	100	+/-100	U	100	P	03/18/2025	14:20	LB135083
	Beryllium	6.00	+/-6.00	U	6.00	P	03/18/2025	14:20	LB135083
	Cadmium	6.00	+/-6.00	U	6.00	P	03/18/2025	14:20	LB135083
	Calcium	2000	+/-2000	U	2000	P	03/18/2025	14:20	LB135083
	Chromium	10.0	+/-10.0	U	10.0	P	03/18/2025	14:20	LB135083
	Cobalt	30.0	+/-30.0	U	30.0	P	03/18/2025	14:20	LB135083
	Copper	20.0	+/-20.0	U	20.0	P	03/18/2025	14:20	LB135083
	Iron	100	+/-100	U	100	P	03/18/2025	14:20	LB135083
	Lead	12.0	+/-12.0	U	12.0	P	03/18/2025	14:20	LB135083

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	G Environmental			SDG No.:	Q1574				
Contract:	GENV01	Lab Code:	CHEM	Case No.:	Q1574	SAS No.:	Q1574		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	03/18/2025	14:20	LB135083
	Manganese	20.0	+/-20.0	U	20.0	P	03/18/2025	14:20	LB135083
	Nickel	40.0	+/-40.0	U	40.0	P	03/18/2025	14:20	LB135083
	Potassium	2000	+/-2000	U	2000	P	03/18/2025	14:20	LB135083
	Selenium	20.0	+/-20.0	U	20.0	P	03/18/2025	14:20	LB135083
	Silver	10.0	+/-10.0	U	10.0	P	03/18/2025	14:20	LB135083
	Sodium	2000	+/-2000	U	2000	P	03/18/2025	14:20	LB135083
	Thallium	40.0	+/-40.0	U	40.0	P	03/18/2025	14:20	LB135083
	Vanadium	40.0	+/-40.0	U	40.0	P	03/18/2025	14:20	LB135083
	Zinc	40.0	+/-40.0	U	40.0	P	03/18/2025	14:20	LB135083
CCB05	Aluminum	100	+/-100	U	100	P	03/18/2025	14:53	LB135083
	Antimony	50.0	+/-50.0	U	50.0	P	03/18/2025	14:53	LB135083
	Arsenic	20.0	+/-20.0	U	20.0	P	03/18/2025	14:53	LB135083
	Barium	100	+/-100	U	100	P	03/18/2025	14:53	LB135083
	Beryllium	6.00	+/-6.00	U	6.00	P	03/18/2025	14:53	LB135083
	Cadmium	6.00	+/-6.00	U	6.00	P	03/18/2025	14:53	LB135083
	Calcium	2000	+/-2000	U	2000	P	03/18/2025	14:53	LB135083
	Chromium	10.0	+/-10.0	U	10.0	P	03/18/2025	14:53	LB135083
	Cobalt	30.0	+/-30.0	U	30.0	P	03/18/2025	14:53	LB135083
	Copper	20.0	+/-20.0	U	20.0	P	03/18/2025	14:53	LB135083
	Iron	100	+/-100	U	100	P	03/18/2025	14:53	LB135083
	Lead	12.0	+/-12.0	U	12.0	P	03/18/2025	14:53	LB135083
	Magnesium	2000	+/-2000	U	2000	P	03/18/2025	14:53	LB135083
	Manganese	20.0	+/-20.0	U	20.0	P	03/18/2025	14:53	LB135083
	Nickel	40.0	+/-40.0	U	40.0	P	03/18/2025	14:53	LB135083
	Potassium	2000	+/-2000	U	2000	P	03/18/2025	14:53	LB135083
	Selenium	20.0	+/-20.0	U	20.0	P	03/18/2025	14:53	LB135083
	Silver	10.0	+/-10.0	U	10.0	P	03/18/2025	14:53	LB135083
	Sodium	2000	+/-2000	U	2000	P	03/18/2025	14:53	LB135083
	Thallium	40.0	+/-40.0	U	40.0	P	03/18/2025	14:53	LB135083
	Vanadium	40.0	+/-40.0	U	40.0	P	03/18/2025	14:53	LB135083
	Zinc	40.0	+/-40.0	U	40.0	P	03/18/2025	14:53	LB135083