

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental **SDG No.:** Q1574
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q1574 **SAS No.:** Q1574

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167161BS							
Arsenic	ug/L	4000	3850		96	80 - 120	P
Barium	ug/L	1000	869		87	80 - 120	P
Cadmium	ug/L	1000	933		93	80 - 120	P
Chromium	ug/L	2000	2010		100	80 - 120	P
Lead	ug/L	5000	4670		93	80 - 120	P
Selenium	ug/L	10000	9550		96	80 - 120	P
Silver	ug/L	380	376		99	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental

SDG No.: Q1574

Contract: GENV01

Lab Code: CHEM

Case No.: Q1574

SAS No.: Q1574

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167166BS Mercury	ug/L	4.0	4.60		115	82 - 119	CV