

Continuing Calibration Report for SequenceID : FF032025AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF015695.D

Aliphatic C9-C12	8283457.000	60.000	3.210	6.833	138057.617	128465.799	-7.466
Aliphatic C12-C16	5675359.000	40.000	6.834	10.268	141883.975	131614.301	-7.803
Aliphatic C16-C21	8726107.000	60.000	10.269	13.635	145435.117	135374.394	-7.432
Aliphatic C21-C28	10964393.000	80.000	13.636	17.299	137054.913	132278.233	-3.611
Aliphatic C28-C40	12363824.000	120.000	17.300	22.199	103031.867	108147.044	4.730
Aliphatic EPH	46013140.000	360.000	3.210	22.199	127814.278	124041.354	-3.042