

Continuing Calibration Report for SequenceID : FF032025AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF015706.D

Aliphatic C9-C12	8315013.000	60.000	3.210	6.833	138583.550	128465.799	-7.876
Aliphatic C12-C16	5730108.000	40.000	6.834	10.268	143252.700	131614.301	-8.843
Aliphatic C16-C21	8783678.000	60.000	10.269	13.635	146394.633	135374.394	-8.141
Aliphatic C21-C28	11052870.000	80.000	13.636	17.299	138160.875	132278.233	-4.447
Aliphatic C28-C40	12850002.000	120.000	17.300	22.199	107083.350	108147.044	0.984
Aliphatic EPH	46731671.000	360.000	3.210	22.199	129810.197	124041.354	-4.651