

Continuing Calibration Report for SequenceID : FF032025AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF015713.D

Aliphatic C9-C12	8210426.000	60.000	3.210	6.833	136840.433	128465.799	-6.519
Aliphatic C12-C16	5637642.000	40.000	6.834	10.268	140941.050	131614.301	-7.086
Aliphatic C16-C21	8671498.000	60.000	10.269	13.635	144524.967	135374.394	-6.759
Aliphatic C21-C28	10915750.000	80.000	13.636	17.299	136446.875	132278.233	-3.151
Aliphatic C28-C40	12634329.000	120.000	17.300	22.199	105286.075	108147.044	2.645
Aliphatic EPH	46069645.000	360.000	3.210	22.199	127971.236	124041.354	-3.168