

Continuing Calibration Report for SequenceID : FG032025AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG015512.D

Aromatic C10-C12	5476572.000	40.000	4.457	6.336	136914.300	133996.317	-2.178
Aromatic C12-C16	8393261.000	60.000	6.337	9.043	139887.683	137151.728	-1.995
Aromatic C16-C21	11290652.000	80.000	9.044	13.372	141133.150	139410.441	-1.236
Aromatic C21-C36	24297615.000	180.000	13.373	18.832	134986.750	130193.774	-3.681
Aromatic EPH	49458100.000	360.000	4.457	18.832	137383.611	133824.086	-2.660