

Continuing Calibration Report for SequenceID : FG032025AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG015530.D

Aromatic C10-C12	5735754.000	40.000	4.457	6.336	143393.850	133996.317	-7.013
Aromatic C12-C16	8839332.000	60.000	6.337	9.043	147322.200	137151.728	-7.415
Aromatic C16-C21	11778156.000	80.000	9.044	13.372	147226.950	139410.441	-5.607
Aromatic C21-C36	24006831.000	180.000	13.373	18.832	133371.283	130193.774	-2.441
Aromatic EPH	50360073.000	360.000	4.457	18.832	139889.092	133824.086	-4.532