

CALIBRATION VERIFICATION SUMMARY

Contract: GENV01

Lab Code: CHEM Case No.: Q1574 SAS No.: Q1574 SDG NO.: Q1574

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL02 Date Analyzed: 03/17/2025

Lab Sample No.: AR1660CCC500 Data File : PO109935.D Time Analyzed: 16:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.787	4.691	4.891	495.420	500.000	-0.9
Aroclor-1016-2	4.806	4.711	4.911	497.010	500.000	-0.6
Aroclor-1016-3	4.862	4.767	4.967	497.830	500.000	-0.4
Aroclor-1016-4	4.983	4.888	5.088	499.000	500.000	-0.2
Aroclor-1016-5	5.240	5.145	5.345	500.320	500.000	0.1
Aroclor-1260-1	6.282	6.188	6.388	468.480	500.000	-6.3
Aroclor-1260-2	6.470	6.377	6.577	466.430	500.000	-6.7
Aroclor-1260-3	6.839	6.746	6.946	449.630	500.000	-10.1
Aroclor-1260-4	7.099	7.005	7.205	446.990	500.000	-10.6
Aroclor-1260-5	7.340	7.247	7.447	446.440	500.000	-10.7
Decachlorobiphenyl	8.747	8.657	8.857	53.600	50.000	7.2
Tetrachloro-m-xylene	3.694	3.598	3.798	47.390	50.000	-5.2

CALIBRATION VERIFICATION SUMMARY

Contract: GENV01

Lab Code: CHEM Case No.: Q1574 SAS No.: Q1574 SDG NO.: Q1574

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL02 Date Analyzed: 03/17/2025

Lab Sample No.: AR1660CCC500 Data File : PO109935.D Time Analyzed: 16:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.773	4.678	4.878	506.990	500.000	1.4
Aroclor-1016-2	4.792	4.697	4.897	518.690	500.000	3.7
Aroclor-1016-3	4.968	4.872	5.072	517.250	500.000	3.5
Aroclor-1016-4	5.010	4.914	5.114	489.360	500.000	-2.1
Aroclor-1016-5	5.223	5.127	5.327	513.620	500.000	2.7
Aroclor-1260-1	6.255	6.161	6.361	474.040	500.000	-5.2
Aroclor-1260-2	6.442	6.348	6.548	476.300	500.000	-4.7
Aroclor-1260-3	6.596	6.502	6.702	469.320	500.000	-6.1
Aroclor-1260-4	7.067	6.974	7.174	450.460	500.000	-9.9
Aroclor-1260-5	7.308	7.214	7.414	446.130	500.000	-10.8
Decachlorobiphenyl	8.697	8.607	8.807	49.430	50.000	-1.1
Tetrachloro-m-xylene	3.692	3.594	3.794	48.820	50.000	-2.4