

### CALIBRATION VERIFICATION SUMMARY

Contract: GENV01

Lab Code: CHEM Case No.: Q1574 SAS No.: Q1574 SDG NO.: Q1574

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL03 Date Analyzed: 03/14/2025

Lab Sample No.: AR1660CCC500 Data File : PP070543.D Time Analyzed: 21:15

| COMPOUND             | RT     | RT WINDOW<br>FROM TO |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|--------|----------------------|--------|--------------------|-------------------|-------|
| Aroclor-1016-1       | 5.676  | 5.577                | 5.777  | 443.660            | 500.000           | -11.3 |
| Aroclor-1016-2       | 5.698  | 5.599                | 5.799  | 480.930            | 500.000           | -3.8  |
| Aroclor-1016-3       | 5.760  | 5.661                | 5.861  | 466.910            | 500.000           | -6.6  |
| Aroclor-1016-4       | 5.858  | 5.758                | 5.958  | 456.460            | 500.000           | -8.7  |
| Aroclor-1016-5       | 6.150  | 6.051                | 6.251  | 453.110            | 500.000           | -9.4  |
| Aroclor-1260-1       | 7.269  | 7.171                | 7.371  | 470.290            | 500.000           | -5.9  |
| Aroclor-1260-2       | 7.523  | 7.425                | 7.625  | 461.520            | 500.000           | -7.7  |
| Aroclor-1260-3       | 7.881  | 7.783                | 7.983  | 466.760            | 500.000           | -6.6  |
| Aroclor-1260-4       | 8.105  | 8.008                | 8.208  | 454.640            | 500.000           | -9.1  |
| Aroclor-1260-5       | 8.425  | 8.328                | 8.528  | 467.430            | 500.000           | -6.5  |
| Decachlorobiphenyl   | 10.243 | 10.152               | 10.352 | 46.560             | 50.000            | -6.9  |
| Tetrachloro-m-xylene | 4.525  | 4.424                | 4.624  | 46.830             | 50.000            | -6.3  |

### CALIBRATION VERIFICATION SUMMARY

Contract: GENV01

Lab Code: CHEM Case No.: Q1574 SAS No.: Q1574 SDG NO.: Q1574

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL03 Date Analyzed: 03/14/2025

Lab Sample No.: AR1660CCC500 Data File : PP070543.D Time Analyzed: 21:15

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| Aroclor-1016-1       | 4.912 | 4.817     | 5.017 | 467.760            | 500.000           | -6.4  |
| Aroclor-1016-2       | 4.932 | 4.836     | 5.036 | 453.880            | 500.000           | -9.2  |
| Aroclor-1016-3       | 5.108 | 5.013     | 5.213 | 459.150            | 500.000           | -8.2  |
| Aroclor-1016-4       | 5.150 | 5.054     | 5.254 | 462.340            | 500.000           | -7.5  |
| Aroclor-1016-5       | 5.365 | 5.270     | 5.470 | 472.460            | 500.000           | -5.5  |
| Aroclor-1260-1       | 6.401 | 6.307     | 6.507 | 458.750            | 500.000           | -8.3  |
| Aroclor-1260-2       | 6.589 | 6.495     | 6.695 | 458.120            | 500.000           | -8.4  |
| Aroclor-1260-3       | 6.743 | 6.649     | 6.849 | 450.000            | 500.000           | -10.0 |
| Aroclor-1260-4       | 7.215 | 7.121     | 7.321 | 441.280            | 500.000           | -11.7 |
| Aroclor-1260-5       | 7.456 | 7.362     | 7.562 | 440.290            | 500.000           | -11.9 |
| Decachlorobiphenyl   | 8.871 | 8.781     | 8.981 | 41.930             | 50.000            | -16.1 |
| Tetrachloro-m-xylene | 3.826 | 3.729     | 3.929 | 48.070             | 50.000            | -3.9  |