

CALIBRATION VERIFICATION SUMMARY

Contract: GENV01

Lab Code: CHEM Case No.: Q1574 SAS No.: Q1574 SDG NO.: Q1574

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL05 Date Analyzed: 03/15/2025

Lab Sample No.: AR1660CCC500 Data File : PP070568.D Time Analyzed: 05:06

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.673	5.577	5.777	464.850	500.000	-7.0
Aroclor-1016-2	5.695	5.599	5.799	501.140	500.000	0.2
Aroclor-1016-3	5.757	5.661	5.861	487.390	500.000	-2.5
Aroclor-1016-4	5.855	5.758	5.958	483.220	500.000	-3.4
Aroclor-1016-5	6.147	6.051	6.251	479.270	500.000	-4.1
Aroclor-1260-1	7.266	7.171	7.371	508.310	500.000	1.7
Aroclor-1260-2	7.519	7.425	7.625	490.450	500.000	-1.9
Aroclor-1260-3	7.878	7.783	7.983	488.320	500.000	-2.3
Aroclor-1260-4	8.102	8.008	8.208	471.720	500.000	-5.7
Aroclor-1260-5	8.422	8.328	8.528	487.240	500.000	-2.6
Decachlorobiphenyl	10.239	10.152	10.352	48.570	50.000	-2.9
Tetrachloro-m-xylene	4.522	4.424	4.624	48.550	50.000	-2.9

CALIBRATION VERIFICATION SUMMARY

Contract: GENV01

Lab Code: CHEM Case No.: Q1574 SAS No.: Q1574 SDG NO.: Q1574

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL05 Date Analyzed: 03/15/2025

Lab Sample No.: AR1660CCC500 Data File : PP070568.D Time Analyzed: 05:06

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.912	4.817	5.017	511.620	500.000	2.3
Aroclor-1016-2	4.930	4.836	5.036	488.190	500.000	-2.4
Aroclor-1016-3	5.107	5.013	5.213	502.650	500.000	0.5
Aroclor-1016-4	5.149	5.054	5.254	495.770	500.000	-0.8
Aroclor-1016-5	5.364	5.270	5.470	540.650	500.000	8.1
Aroclor-1260-1	6.400	6.307	6.507	493.180	500.000	-1.4
Aroclor-1260-2	6.587	6.495	6.695	483.690	500.000	-3.3
Aroclor-1260-3	6.741	6.649	6.849	471.010	500.000	-5.8
Aroclor-1260-4	7.213	7.121	7.321	458.170	500.000	-8.4
Aroclor-1260-5	7.454	7.362	7.562	453.230	500.000	-9.4
Decachlorobiphenyl	8.869	8.781	8.981	46.240	50.000	-7.5
Tetrachloro-m-xylene	3.826	3.729	3.929	49.820	50.000	-0.4